

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

LAB CHRONICLE

OrderID:	Q1698	OrderDate:	4/2/2025 12:14:00 PM
Client:	Tully Construction Co., Inc.	Project:	MTA Rockaway Park
Contact:	Dean Devoe	Location:	I41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1698-02	B-9	SOIL			04/02/25 08:10			04/02/25
			Corrosivity	9045D			04/02/25 16:30	
			Ignitability	1030			04/03/25 11:05	
			Reactive Cyanide	9012B		04/04/25	04/04/25 13:16	
			Reactive Sulfide	9034		04/03/25	04/03/25 12:42	
Q1698-08	B-10	SOIL			04/02/25 08:35			04/02/25
			Corrosivity	9045D			04/02/25 16:35	
			Ignitability	1030			04/03/25 11:12	
			Reactive Cyanide	9012B		04/04/25	04/04/25 13:23	
			Reactive Sulfide	9034		04/03/25	04/03/25 12:44	



SAMPLE DATA

Report of Analysis

Client:	Tully Construction Co., Inc.	Date Collected:	04/02/25 08:10
Project:	MTA Rockaway Park	Date Received:	04/02/25
Client Sample ID:	B-9	SDG No.:	Q1698
Lab Sample ID:	Q1698-02	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	10.6	H	1	0	0	pH		04/02/25 16:30	9045D
Ignitability	NO		1	0	0	oC		04/03/25 11:05	1030
Reactive Cyanide	0.0084	U	1	0.0084	0.050	mg/Kg	04/04/25 10:00	04/04/25 13:16	9012B
Reactive Sulfide	3.17	J	1	0.20	10.0	mg/Kg	04/03/25 08:50	04/03/25 12:42	9034

Comments: pH result reported at temperature 24.6 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tully Construction Co., Inc.	Date Collected:	04/02/25 08:35
Project:	MTA Rockaway Park	Date Received:	04/02/25
Client Sample ID:	B-10	SDG No.:	Q1698
Lab Sample ID:	Q1698-08	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	10.6	H	1	0	0	pH		04/02/25 16:35	9045D
Ignitability	NO		1	0	0	oC		04/03/25 11:12	1030
Reactive Cyanide	0.0084	U	1	0.0084	0.050	mg/Kg	04/04/25 10:00	04/04/25 13:23	9012B
Reactive Sulfide	4.78	J	1	0.20	10.0	mg/Kg	04/03/25 08:50	04/03/25 12:44	9034

Comments: pH result reported at temperature 23.7 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Project: MTA Rockaway Park

RunNo.: LB135276

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	04/02/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	04/02/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	04/02/2025

Initial and Continuing Calibration Verification

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Project: MTA Rockaway Park

RunNo.: LB135312

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.094	0.099	95	85-115	04/04/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/04/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/04/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/04/2025

Initial and Continuing Calibration Verification

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Project: MTA Rockaway Park

RunNo.: LB135312

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------

Initial and Continuing Calibration Blank Summary

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Project: MTA Rockaway Park

RunNo.: LB135312

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/04/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/04/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/04/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/04/2025

Initial and Continuing Calibration Blank Summary

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Project: MTA Rockaway Park

RunNo.: LB135312

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------

Preparation Blank Summary

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Project: MTA Rockaway Park

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: PB167444BL							
Reactive Sulfide	mg/Kg	< 5.0000	5.0000	U	0.201	10	04/03/2025
Sample ID: PB167462BL							
Reactive Cyanide	mg/Kg	< 0.0250	0.0250	U	0.0084	0.05	04/04/2025

Matrix Spike Summary

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Project:

Sample ID:

Client ID:

Percent Solids for Spike Sample:

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
---------	-------	------------------------	------------------	--------------------	------------------	--------------------	----------------	--------------------	----------	------	------------------

Duplicate Sample Summary

Client: Tully Construction Co., Inc. Project: MTA Rockaway Park Client ID: TP-4DUP	SDG No.: Q1698 Sample ID: Q1680-04 Percent Solids for Spike Sample: 100
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Reactive Sulfide	mg/Kg	+/-20	4.74	J	4.74	J	1	0		04/03/2025
Reactive Cyanide	mg/Kg	+/-20	0.0084	U	0.0084	U	1	0		04/04/2025

Duplicate Sample Summary

Client:	Tully Construction Co., Inc.	SDG No.:	Q1698
Project:	MTA Rockaway Park	Sample ID:	Q1694-01
Client ID:	TP-18DUP	Percent Solids for Spike Sample:	88.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		04/03/2025

Duplicate Sample Summary

Client: Tully Construction Co., Inc.	SDG No.: Q1698
Project: MTA Rockaway Park	Sample ID: Q1694-04
Client ID: TP-18DUP	Percent Solids for Spike Sample: 100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Corrosivity	pH	+/-20	7.74		7.75		1	0.13		04/02/2025



Client: Tully Construction Co., Inc.

SDG No.: Q1698

Project:

Run No.:

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID								



RAW DATA

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135276

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/02/2025	15:30
2	CAL2	1	Water	NA	NA	20.2	7.00	04/02/2025	15:31
3	CAL3	1	Water	NA	NA	20.3	10.02	04/02/2025	15:35
4	ICV	1	Water	NA	NA	20.2	7.01	04/02/2025	15:38
5	CCV1	1	Water	NA	NA	20.3	2.01	04/02/2025	15:44
6	Q1694-04	1	Solid	20.02	20	22.6	7.74	04/02/2025	15:50
7	Q1694-04DUP	1	Solid	20.03	20	22.7	7.75	04/02/2025	15:51
8	Q1695-04	1	Solid	20.04	20	22.8	9.67	04/02/2025	16:10
9	Q1695-08	1	Solid	20.02	20	22.6	8.70	04/02/2025	16:20
10	Q1698-02	1	Solid	20.05	20	24.6	10.55	04/02/2025	16:30
11	Q1698-08	1	Solid	20.04	20	23.7	10.57	04/02/2025	16:35
12	CCV2	1	Water	NA	NA	20.2	12.02	04/02/2025	16:40

WORKLIST(Hardcopy Internal Chain)

17135276

WorkList Name : corrosivity q1695 WorkList ID : 188684 Department : Wet-Chemistry Date : 04-02-2025 14:38:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1694-04	TP-18	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	04/01/2025	9045D
Q1695-04	TT-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	04/02/2025	9045D
Q1695-08	TT-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	04/02/2025	9045D
Q1698-02	B-9	Solid	Corrosivity	Cool 4 deg C	TULL02	I41	04/02/2025	9045D
Q1698-08	B-10	Solid	Corrosivity	Cool 4 deg C	TULL02	I41	04/02/2025	9045D

Date/Time 04/02/25 14:55

Raw Sample Received by: JB CWC

Raw Sample Relinquished by: JB CWC

Date/Time 04/02/25

Raw Sample Received by: JB CWC

Raw Sample Relinquished by: JB CWC

18.00

JB CWC

JB CWC

Analytical Summary Report

Analysis Method: 1030
Parameter: Ignitability
Run Number: LB135281

Reviewed By: rubina

Supervisor Review By: Iwona

Seq	LabID	ClientID	DF	matrix	Result Status	Burning Rate	Anal Date	Anal Time
1	Q1694-01	TP-18	1	Solid	NO	0.00	04/03/2025	10:00
2	Q1694-01DUP	TP-18DUP	1	Solid	NO	0.00	04/03/2025	10:07
3	Q1694-04	TP-18	1	Solid	NO	0.00	04/03/2025	10:15
4	Q1695-01	TT-2	1	Solid	NO	0.00	04/03/2025	10:35
5	Q1695-04	TT-2	1	Solid	NO	0.00	04/03/2025	10:42
6	Q1695-05	TT-3	1	Solid	NO	0.00	04/03/2025	10:50
7	Q1695-08	TT-3	1	Solid	NO	0.00	04/03/2025	10:57
8	Q1698-02	B-9	1	Solid	NO	0.00	04/03/2025	11:05
9	Q1698-08	B-10	1	Solid	NO	0.00	04/03/2025	11:12

Burning Rate = $\frac{\text{Length (mm)}}{\text{Total Time (sec)}}$

WORKLIST(Hardcopy Internal Chain)

66135281

WorkList Name : ign-4-3

WorkList ID : 188699

Department : Wet-Chemistry

Date : 04-03-2025 08:17:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1694-01 B	TP-18	Solid	Ignitability	Cool 4 deg C	PSEG03	I41	04/01/2025	1030
Q1694-04 B	TP-18	Solid	Ignitability	Cool 4 deg C	PSEG03	I41	04/01/2025	1030
Q1695-01 C	TT-2	Solid	Ignitability	Cool 4 deg C	PSEG03	I41	04/02/2025	1030
Q1695-04 C	TT-2	Solid	Ignitability	Cool 4 deg C	PSEG03	I41	04/02/2025	1030
Q1695-05 B	TT-3	Solid	Ignitability	Cool 4 deg C	PSEG03	I41	04/02/2025	1030
Q1695-08 B	TT-3	Solid	Ignitability	Cool 4 deg C	PSEG03	I41	04/02/2025	1030
Q1698-02 A	B-9	Solid	Ignitability	Cool 4 deg C	TULL02	I41	04/02/2025	1030
Q1698-08 B	B-10	Solid	Ignitability	Cool 4 deg C	TULL02	I41	04/02/2025	1030

Date/Time 04/03/2025 08:20
 Raw Sample Received by: RM CWO
 Raw Sample Relinquished by: 28 CWO

Date/Time 04/03/2025 11:30
 Raw Sample Received by: 28 CWO
 Raw Sample Relinquished by: RM CWO

Analysis Method: 9034

ANALYST: rubina

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB135282

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3114
Starch Solution, 4L	W3149

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB167444BL		1	5.00	50	2.00	0.00	1.94	1.94	0.06	0.00	0.00	04/03/2025	12:15
2	Q1680-04		1	5.06	50	2.00	0.00	1.88	1.88	0.12	0.06	4.74	04/03/2025	12:18
3	Q1680-04DUP		1	5.06	50	2.00	0.00	1.88	1.88	0.12	0.06	4.74	04/03/2025	12:21
4	Q1681-04		1	5.03	50	2.00	0.00	1.90	1.90	0.10	0.04	3.18	04/03/2025	12:24
5	Q1693-04		1	5.07	50	2.00	0.00	1.88	1.88	0.12	0.06	4.73	04/03/2025	12:27
6	Q1693-08		1	5.04	50	2.00	0.00	1.90	1.90	0.10	0.04	3.17	04/03/2025	12:30
7	Q1693-12		1	5.01	50	2.00	0.00	1.86	1.86	0.14	0.08	6.39	04/03/2025	12:33
8	Q1694-04		1	5.04	50	2.00	0.00	1.86	1.86	0.14	0.08	6.35	04/03/2025	12:35
9	Q1695-04		1	5.06	50	2.00	0.00	1.88	1.88	0.12	0.06	4.74	04/03/2025	13:37
10	Q1695-08		1	5.07	50	2.00	0.00	1.90	1.90	0.10	0.04	3.16	04/03/2025	12:40
11	Q1698-02		1	5.05	50	2.00	0.00	1.90	1.90	0.10	0.04	3.17	04/03/2025	12:42
12	Q1698-08		1	5.02	50	2.00	0.00	1.88	1.88	0.12	0.06	4.78	04/03/2025	12:44

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume

Page: 1

Reviewed by : NF Instrument ID : Konelab

Reviewed by : NF Instrument ID : Konelab

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.253	0.0	0.073	
ICB1	0.683	0.0	0.001	
CCV1	249.105	0.0	0.193	
CCB1	0.025	0.0	0.000	
PB167462BL	0.167	0.0	0.000	
Q1680-04	0.125	0.0	0.000	
Q1680-04DUP	0.105	0.0	0.000	
Q1681-04	0.277	0.0	0.000	
Q1693-04	0.234	0.0	0.000	
Q1693-08	0.384	0.0	0.000	
Q1693-12	0.580	0.0	0.000	
Q1694-04	0.359	0.0	0.000	
Q1695-04	0.322	0.0	0.000	
Q1695-08	0.447	0.0	0.000	
CCV2	250.630	0.0	0.194	
CCB2	0.293	0.0	0.000	
Q1698-02	0.168	0.0	0.000	
Q1698-08	0.249	0.0	0.000	
Q1712-04	0.350	0.0	0.000	
Q1712-08	0.331	0.0	0.000	
CCV3	251.847	0.0	0.195	
CCB3	0.370	0.0	0.000	

N	22
Mean	38.696
SD	88.4361
CV%	228.54

Aquakem v. 7.2AQ1

Results from time period:

Fri Apr 04 12:56:18 2025

Fri Apr 04 13:27:40 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time Stat
0.0PPBCN	A	Total CN	P	0.1799	µg/l	4/4/2025 10:29:29
5.0PPBCN	A	Total CN	P	5.4701	µg/l	4/4/2025 10:29:30
10PPBCN	A	Total CN	P	10.2111	µg/l	4/4/2025 10:29:31
50PPBCN	A	Total CN	P	49.8763	µg/l	4/4/2025 10:29:32
100PPBCN	A	Total CN	P	98.7893	µg/l	4/4/2025 10:29:33
250PPBCN	A	Total CN	P	250.4552	µg/l	4/4/2025 10:29:34
500PPBCN	A	Total CN	P	500.0179	µg/l	4/4/2025 10:29:35
ICV1	S	Total CN	P	94.2535	µg/l	4/4/2025 12:56:19
ICB1	S	Total CN	P	0.6831	µg/l	4/4/2025 12:56:20
CCV1	S	Total CN	P	249.1047	µg/l	4/4/2025 12:56:23
CCB1	S	Total CN	P	0.0253	µg/l	4/4/2025 12:56:24
PB167462BL	S	Total CN	P	0.1666	µg/l	4/4/2025 12:56:26
Q1680-04	S	Total CN	P	0.1253	µg/l	4/4/2025 13:03:50
Q1680-04DUP	S	Total CN	P	0.1053	µg/l	4/4/2025 13:03:51
Q1681-04	S	Total CN	P	0.2767	µg/l	4/4/2025 13:03:54
Q1693-04	S	Total CN	P	0.2336	µg/l	4/4/2025 13:03:55
Q1693-08	S	Total CN	P	0.3844	µg/l	4/4/2025 13:03:57
Q1693-12	S	Total CN	P	0.5801	µg/l	4/4/2025 13:03:59
Q1694-04	S	Total CN	P	0.3589	µg/l	4/4/2025 13:11:29
Q1695-04	S	Total CN	P	0.3215	µg/l	4/4/2025 13:11:31
Q1695-08	S	Total CN	P	0.4472	µg/l	4/4/2025 13:11:34
CCV2	S	Total CN	P	250.6301	µg/l	4/4/2025 13:16:12
CCB2	S	Total CN	P	0.2934	µg/l	4/4/2025 13:16:14
Q1698-02	S	Total CN	P	0.1675	µg/l	4/4/2025 13:16:15
Q1698-08	S	Total CN	P	0.2491	µg/l	4/4/2025 13:23:48
Q1712-04	S	Total CN	P	0.3495	µg/l	4/4/2025 13:23:49
Q1712-08	S	Total CN	P	0.3314	µg/l	4/4/2025 13:23:52
CCV3	S	Total CN	P	251.8471	µg/l	4/4/2025 13:27:38
CCB3	S	Total CN	P	0.3697	µg/l	4/4/2025 13:27:40

=====

Calibration results	Aquakem 7.2AQ1	Page: 1
---------------------	----------------	---------

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : NF Instrument ID : Konelab

4/4/2025 10:33


Test Total CN

Accepted 4/4/2025 10:33

Factor 1292
Bias 0

Coeff. of det. 0.999990

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.0PPBCN	0.000	0.1799	0.0000	
2	5.0PPBCN	0.004	5.4701	5.0000	9.4
3	10PPBCN	0.008	10.2111	10.0000	2.1
4	50PPBCN	0.039	49.8763	50.0000	-0.2
5	100PPBCN	0.076	98.7893	100.0000	-1.2
6	250PPBCN	0.194	250.4552	250.0000	0.2
7	500PPBCN	0.387	500.0179	500.0000	0.0

NF
04.04.2025

SOP ID : M9030B-Sulfide-12

SDG No : N/A

Matrix : SOIL

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#1

Block ID : MC-1,MC-2

Weigh By : RM

Start Digest Date: 04/03/2025 Time : 08:50 Temp : N/A

End Digest Date: 04/03/2025 Time : 10:20 Temp : N/A

Digestion tube ID : M5595

Filter paper ID : N/A

pH Meter ID : N/A

Block Thermometer ID : N/A

Prep Technician Signature: RM

Supervisor Signature: 12

Standard Name	MLS USED	STD REF. # FROM LOG
PBS003	50.0ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.5M ZINC ACETATE	5.0ML	WP111004
FORMALDEHYDE	2.0ML	W2725
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

04/03/2025 RM

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB167444BL	PBS444	5.00	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1680-04DUP	TP-4DUP	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1680-04	TP-4	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1681-04	TP-12	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1693-04	WCS-TP-3	5.07	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1693-08	WCS-TP-2	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1693-12	WCS-TP-1	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1694-04	TP-18	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1695-04	TT-2	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1695-08	TT-3	5.07	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1698-02	B-9	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1698-08	B-10	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A

SOP ID : M9012B-Total, Amenable and Reactive Cyanide-20

SDG No : N/A

Start Digest Date: 04/04/2025 **Time :** 10:00 **Temp :** N/A

Matrix : SOIL

End Digest Date: 04/04/2025 **Time :** 11:30 **Temp :** N/A

Pipette ID : N/A

Balance ID : WC SC-7

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : MC-1, MC-2

Filter paper ID : N/A

Prep Technician Signature: *NF*

Weigh By : NF

pH Meter ID : N/A

Supervisor Signature: *12*

Standard Name	MLS USED	STD REF. # FROM LOG
PBS003	50ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.25N NaOH	50.0ML	WP111294
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Comment

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04.04.2025, 11:50	NF/WC	NF/WC
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB167462BL	PBS462	5.00	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1680-04DUP	TP-4DUP	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1680-04	TP-4	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1681-04	TP-12	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1693-04	WCS-TP-3	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1693-08	WCS-TP-2	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1693-12	WCS-TP-1	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1694-04	TP-18	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1695-04	TT-2	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1695-08	TT-3	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1698-02	B-9	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1698-08	B-10	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1712-04	Z-05A	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q1712-08	TT-7	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A

Instrument ID: WC PH METER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135276

Review By	jignesh	Review On	4/3/2025 8:51:32 AM
Supervise By	Iwona	Supervise On	4/3/2025 11:37:35 AM
SubDirectory	LB135276	Test	Corrosivity
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3178,W3093,W3191,W3071,W3161,W3072		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	04/02/25 15:30		Jignesh	OK
2	CAL2	CAL2	CAL	04/02/25 15:31		Jignesh	OK
3	CAL3	CAL3	CAL	04/02/25 15:35		Jignesh	OK
4	ICV	ICV	ICV	04/02/25 15:38		Jignesh	OK
5	CCV1	CCV1	CCV	04/02/25 15:44		Jignesh	OK
6	Q1694-04	TP-18	SAM	04/02/25 15:50		Jignesh	OK
7	Q1694-04DUP	TP-18DUP	DUP	04/02/25 15:51		Jignesh	OK
8	Q1695-04	TT-2	SAM	04/02/25 16:10		Jignesh	OK
9	Q1695-08	TT-3	SAM	04/02/25 16:20		Jignesh	OK
10	Q1698-02	B-9	SAM	04/02/25 16:30		Jignesh	OK
11	Q1698-08	B-10	SAM	04/02/25 16:35		Jignesh	OK
12	CCV2	CCV2	CCV	04/02/25 16:40		Jignesh	OK

Instrument ID: FLAME

Daily Analysis Runlog For Sequence/QC Batch ID # LB135281

Review By	rubina	Review On	4/3/2025 12:57:35 PM
Supervise By	Iwona	Supervise On	4/3/2025 1:02:43 PM
SubDirectory	LB135281	Test	Ignitability
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	Q1694-01	TP-18	SAM	04/03/25 10:00		rubina	OK
2	Q1694-01DUP	TP-18DUP	DUP	04/03/25 10:07		rubina	OK
3	Q1694-04	TP-18	SAM	04/03/25 10:15		rubina	OK
4	Q1695-01	TT-2	SAM	04/03/25 10:35		rubina	OK
5	Q1695-04	TT-2	SAM	04/03/25 10:42		rubina	OK
6	Q1695-05	TT-3	SAM	04/03/25 10:50		rubina	OK
7	Q1695-08	TT-3	SAM	04/03/25 10:57		rubina	OK
8	Q1698-02	B-9	SAM	04/03/25 11:05		rubina	OK
9	Q1698-08	B-10	SAM	04/03/25 11:12		rubina	OK

Instrument ID: TITRAMETRIC

Daily Analysis Runlog For Sequence/QC Batch ID # LB135282

Review By	rubina	Review On	4/3/2025 12:57:49 PM
Supervise By	Iwona	Supervise On	4/3/2025 1:02:29 PM
SubDirectory	LB135282	Test	Reactive Sulfide
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3105,W3114,W3149		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	PB167444BL	PB167444BL	MB	04/03/25 12:15		rubina	OK
2	Q1680-04	TP-4	SAM	04/03/25 12:18		rubina	OK
3	Q1680-04DUP	TP-4DUP	DUP	04/03/25 12:21		rubina	OK
4	Q1681-04	TP-12	SAM	04/03/25 12:24		rubina	OK
5	Q1693-04	WCS-TP-3	SAM	04/03/25 12:27		rubina	OK
6	Q1693-08	WCS-TP-2	SAM	04/03/25 12:30		rubina	OK
7	Q1693-12	WCS-TP-1	SAM	04/03/25 12:33		rubina	OK
8	Q1694-04	TP-18	SAM	04/03/25 12:35		rubina	OK
9	Q1695-08	TT-3	SAM	04/03/25 12:40		rubina	OK
10	Q1698-02	B-9	SAM	04/03/25 12:42		rubina	OK
11	Q1698-08	B-10	SAM	04/03/25 12:44		rubina	OK
12	Q1695-04	TT-2	SAM	04/03/25 13:37		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB135312

Review By	Review On		
Supervise By	Supervise On		
SubDirectory	LB135312	Test	Reactive Cyanide

STD. NAME	STD REF.#
ICAL Standard	WP112597,WP112598,WP112599,WP112600,WP112601,WP112602,WP112603
ICV Standard	WP112605
CCV Standard	WP112598
ICSA Standard	N/A
CRI Standard	N/A
LCS Standard	N/A
Chk Standard	WP111035,WP110103,WP112604

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPBCN	0.0PPBCN	CAL1	04/04/25 10:29		Niha	OK
2	5.0PPBCN	5.0PPBCN	CAL2	04/04/25 10:29		Niha	OK
3	10PPBCN	10PPBCN	CAL3	04/04/25 10:29		Niha	OK
4	50PPBCN	50PPBCN	CAL4	04/04/25 10:29		Niha	OK
5	100PPBCN	100PPBCN	CAL5	04/04/25 10:29		Niha	OK
6	250PPBCN	250PPBCN	CAL6	04/04/25 10:29		Niha	OK
7	500PPBCN	500PPBCN	CAL7	04/04/25 10:29		Niha	OK
8	ICV1	ICV1	ICV	04/04/25 12:56		Niha	OK
9	ICB1	ICB1	ICB	04/04/25 12:56		Niha	OK
10	CCV1	CCV1	CCV	04/04/25 12:56		Niha	OK
11	CCB1	CCB1	CCB	04/04/25 12:56		Niha	OK
12	PB167462BL	PB167462BL	MB	04/04/25 12:56		Niha	OK
13	Q1680-04	TP-4	SAM	04/04/25 13:03		Niha	OK
14	Q1680-04DUP	TP-4DUP	DUP	04/04/25 13:03		Niha	OK
15	Q1681-04	TP-12	SAM	04/04/25 13:03		Niha	OK
16	Q1693-04	WCS-TP-3	SAM	04/04/25 13:03		Niha	OK
17	Q1693-08	WCS-TP-2	SAM	04/04/25 13:03		Niha	OK
18	Q1693-12	WCS-TP-1	SAM	04/04/25 13:03		Niha	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB135312

Review By	Review On
Supervise By	Supervise On
SubDirectory LB135312	Test Reactive Cyanide
STD. NAME	STD REF.#
ICAL Standard	WP112597,WP112598,WP112599,WP112600,WP112601,WP112602,WP112603
ICV Standard	WP112605
CCV Standard	WP112598
ICSA Standard	N/A
CRI Standard	N/A
LCS Standard	N/A
Chk Standard	WP111035,WP110103,WP112604

19	Q1694-04	TP-18	SAM	04/04/25 13:11		Niha	OK
20	Q1695-04	TT-2	SAM	04/04/25 13:11		Niha	OK
21	Q1695-08	TT-3	SAM	04/04/25 13:11		Niha	OK
22	CCV2	CCV2	CCV	04/04/25 13:16		Niha	OK
23	CCB2	CCB2	CCB	04/04/25 13:16		Niha	OK
24	Q1698-02	B-9	SAM	04/04/25 13:16		Niha	OK
25	Q1698-08	B-10	SAM	04/04/25 13:23		Niha	OK
26	Q1712-04	Z-05A	SAM	04/04/25 13:23		Niha	OK
27	Q1712-08	TT-7	SAM	04/04/25 13:23		Niha	OK
28	CCV3	CCV3	CCV	04/04/25 13:27		Niha	OK
29	CCB3	CCB3	CCB	04/04/25 13:27		Niha	OK

Prep Standard - Chemical Standard Summary

Order ID : Q1698

Test : Corrosivity,Ignitability,Percent Solids,Reactive Cyanide,Reactive Sulfide

Prepbatch ID : PB167444,PB167462,

Sequence ID/Qc Batch ID: LB135276,LB135281,LB135282,LB135312,

Standard ID :

WP110103,WP111004,WP111035,WP111294,WP111296,WP112596,WP112597,WP112598,WP112599,WP112600,WP112601,WP112602,WP112603,WP112604,WP112605,

Chemical ID :

M6121,W2668,W2725,W2882,W2926,W3019,W3071,W3072,W3093,W3105,W3112,W3113,W3114,W3138,W3139,W3149,W3154,W3161,W3178,W3191,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
539	CN BUFFER	WP110103	10/08/2024	04/08/2025	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych 10/08/2024
<u>FROM</u> 138.00000gram of W2668 + 862.00000ml of W3112 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
160	0.5M ZINC ACETATE	WP111004	12/09/2024	05/13/2025	Rubina Mughal	WETCHEM_SCALE_8 (WC SC-7)	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 12/09/2024
FROM 0.88900L of W3112 + 1.00000ml of M6121 + 110.00000gram of W2926 = Final Quantity: 1000.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
607	PYRIDINE-BARBITURIC ACID	WP111035	12/09/2024	04/30/2025	Niha Farheen Shaik	WETCHEM_S CALE_5 (WC SC-5)	Glass Pipette-A	Iwona Zarych 12/10/2024
<u>FROM</u>	145.00000ml of W3112 + 15.00000gram of W2882 + 15.00000ml of M6121 + 75.00000ml of W3019 = Final Quantity: 250.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
11	Sodium hydroxide absorbing solution 0.25 N	WP111294	01/07/2025	07/07/2025	Niha Farheen Shaik	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych 01/07/2025
<u>FROM</u> 21.00000L of W3112 + 210.00000gram of W3113 = Final Quantity: 21.000 L								

[illegible]

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3456	Cyanide Intermediate Working Std, 5PPM	WP112596	04/04/2025	04/05/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 04/04/2025
<u>FROM</u> 0.25000ml of W3154 + 49.75000ml of WP111294 = Final Quantity: 50.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4	Calibration standard 500 ppb	WP112597	04/04/2025	04/05/2025	Niha Farheen Shaik	None	Glass Pipette-A	Iwona Zarych 04/04/2025

FROM 45.00000ml of WP111294 + 5.00000ml of WP112596 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3761	Calibration-CCV CN Standard 250 ppb	WP112598	04/04/2025	04/05/2025	Niha Farheen Shaik	None	Glass Pipette-A	Iwona Zarych 04/04/2025

FROM 2.50000ml of WP112596 + 47.50000ml of WP111294 = Final Quantity: 50.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
6	Calibration Standard 100 ppb	WP112599	04/04/2025	04/05/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 04/04/2025
<u>FROM</u>	1.00000ml of WP112596 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
7	Calibration Standard 50 ppb	WP112600	04/04/2025	04/05/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 04/04/2025
<u>FROM</u>	0.50000ml of WP112596 + 49.50000ml of WP111294 = Final Quantity: 50.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
8	Calibration Standard 10 ppb	WP112601	04/04/2025	04/05/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3	Iwona Zarych 04/04/2025
<u>FROM</u>	1.00000ml of WP112597 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
9	Calibration Standard 5 ppb	WP112602	04/04/2025	04/05/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 04/04/2025
<u>FROM</u>	0.50000ml of WP112597 + 49.50000ml of WP111294 = Final Quantity: 50.000 ml							

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
167	0 ppb CN calibration std	WP112603	04/04/2025	04/05/2025	Niha Farheen Shaik	None	None	Iwona Zarych
								04/04/2025

FROM 50.00000ml of WP111294 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1582	Chloramine T solution, 0.014M	WP112604	04/04/2025	04/05/2025	Niha Farheen Shaik	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych
								04/04/2025

FROM 0.08000gram of W3139 + 20.00000ml of W3112 = Final Quantity: 20.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2168	RCN ICV STD, 100 PPB	WP112605	04/04/2025	04/05/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 04/04/2025
<u>FROM</u>	1.00000ml of WP111296 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	0000275677	05/13/2025	11/13/2024 / Eman	10/13/2024 / Eman	M6121

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3818-5 / SODIUM PHOSPHATE, MONOBAS/HYD, CRYST, ACS, 2.5 KG	0000225799	12/03/2025	04/05/2021 / Alexander	02/10/2020 / apatel	W2668

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EMD-FX0410-5 / FORMALDEHYDE SOLUTION 450ML	60045	06/22/2025	08/19/2024 / lwona	06/22/2020 / apatel	W2725

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EM-BX0035-3 / Barbituric Acid, 100 gms	1.00132.0100	04/30/2025	12/07/2021 /	11/30/2021 / apatel	W2882

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J4296-1 / ZINC ACETATE, DIHYD, CRYST, ACS, 500G	383058	07/05/2027	07/05/2022 / ketankumar	07/05/2022 / ketankumar	W2926

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
SIGMA ALDRICH	270970-1L / Pyridine 1L	SHBQ2113	04/03/2028	04/03/2023 / lwona	04/03/2023 / lwona	W3019

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14455-3 / buffer solution pH 7 yellow	4308H30	07/31/2025	01/02/2024 / JIGNESH	12/06/2023 / Iwona	W3071

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14940-1 / Buffer Solution, PH12 (500ml)	2310P21	04/30/2025	01/02/2024 / JIGNESH	12/07/2023 / Iwona	W3072

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	566002 / BUFFER PH 7.00 GREEN 1PINT PK6	44001f99	12/31/2025	04/03/2024 / jignesh	04/02/2024 / jignesh	W3093

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL69870-8 / SODIUM THIOSULFATE,0.025N,4LIT RE	4403S13	09/30/2025	04/22/2024 / Iwona	04/22/2024 / Iwona	W3105

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-7 / Sodium Hydroxide Pellets 12 Kg	23B1556310	12/31/2025	07/08/2024 / Iwona	07/08/2024 / Iwona	W3113

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL35830-4 / IODINE SOLUTION .025N 1L	2405D89	05/31/2025	07/10/2024 / lwona	07/10/2024 / lwona	W3114

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	LC135457 / Cyanide Standard, 1000 PPM, Second Source	44080060	01/30/2025	09/06/2024 / lwona	08/28/2024 / lwona	W3138

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	JTE494-6 / CHLORAMINE-T BAKER 250GM	10239484	09/09/2029	09/09/2024 / lwona	09/09/2024 / lwona	W3139

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL70850-8 / Starch Solution, 4L	4408P62	08/31/2026	10/16/2024 / lwona	10/16/2024 / lwona	W3149

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	RC2543-4 / CYANIDE STD 1000PPM 4OZ	1411J58	05/31/2025	12/02/2024 / lwona	12/02/2024 / lwona	W3154

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL13850-1 / Buffer Solution, PH2 (500ml)	2411E26	10/31/2026	12/09/2024 / lwona	12/09/2024 / lwona	W3161

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14055-3 / PH 4 BUFFER SOLUTION	2411A93	10/30/2026	04/01/2025 / JIGNESH	01/27/2025 / jignesh	W3178

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	1601-1 / PH 10.01 BUFFER,COLOR CD 475ML	2410F80	03/31/2026	04/01/2025 / JIGNESH	03/13/2025 / jignesh	W3191


W3071
Rec 12/6/23

Certificate of Analysis 12

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4308H30

Product Number: 1551

Manufacture Date: AUG 09, 2023

Expiration Date: JUL 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Yellow liquid	Passed

*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.002	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (08/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

W3019
rec 4/3/23

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.comEmail USA: techserv@sial.comOutside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:

Pyridine - anhydrous, 99.8%

Product Number:

270970

Batch Number:

SHBQ2113

Brand:

SIAL

CAS Number:

110-86-1

MDL Number:

MFCD00011732

Formula:


C₅H₅N

Formula Weight:

79.10 g/mol

Quality Release Date:

15 DEC 2022



Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Form)	Liquid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Purity (GC)	≥ 99.75 %	99.99 %
Water (by Karl Fischer)	≤ 0.003 %	0.002 %
Residue on Evaporation	≤ 0.0005 %	< 0.0001 %


Larry Coers, Director
Quality Control
Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





RICCA CHEMICAL COMPANY®

1841 Broad Street
Pocomoke City, MD 21851
<http://www.riccachemical.com>
1-888-GO-RICCA
customerservice@riccachemical.com

W 3072
REC. 12/01/23
12

Certificate of Analysis

Buffer, Reference Standard, pH 12.00 ± 0.01 at 25°C

Lot Number: 2310P21

Product Number: 1615

Manufacture Date: OCT 24, 2023

Expiration Date: APR 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C	15	20	25	30	35	40
pH	12.35	12.17	11.99	11.78	11.62	11.46

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed


*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	12.005	0.02	186-I-g, 186-II-g, 191d

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1615-1	4 L natural poly	18 months
1615-16	500 mL clear PET-G	18 months
1615-32	1 L natural poly	18 months
1615-5	20 L Cubitainer®	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Sharon Travers (10/24/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Certificate of Analysis

Date of Release: 2/26/2020

Name: Formaldehyde Solution
GR ACS
Meets ACS Specifications

Item No: FX0410 all size codes

Lot / Batch No: 60045

Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Assay	36.5	38.0	36.71	%
Chloride (Cl)		5	<5	ppm
Color (APHA)		10	<10	
Form			Passes test	
Heavy metals (as Pb)		5	<5	ppm
Iron (Fe)		5	0.6	ppm
Residue after ignition		0.005	<0.0050	%
Sulfate (SO ₄)		0.002	<0.0020	%
Titration acid		0.006	<0.0060	meq/g

Heather Sinn,

Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Millipore Corporation, an affiliate of Merck KGaA, Darmstadt, Germany
290 Concord Road
Billerica, MA 01821
U.S.A

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

Hydrochloric Acid, 36.5-38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

avantor™



R → 16/13/24
Met dig

M 6121

Material No.: 9530-33
Batch No.: 0000275677
Manufactured Date: 2020/12/16
Retest Date: 2025/12/15
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS - Assay (as HCl) (by acid-base titrn)	36.5 - 38.0 %	37.6
ACS - Color (APHA)	<= 10	5
ACS - Residue after Ignition	<= 3 ppm	1
ACS - Specific Gravity at 60°/60°F	1.185 - 1.192	1.190
ACS - Bromide (Br)	<= 0.005 %	< 0.005
ACS - Extractable Organic Substances	<= 5 ppm	1
ACS - Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities - Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities - Aluminum (Al)	<= 10.0 ppb	< 0.2
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities - Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities - Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities - Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities - Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities - Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities - Calcium (Ca)	<= 50.0 ppb	29.7
Trace Impurities - Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities - Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities - Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities - Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Material No.: 9530-33
Batch No.: 0000275677

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	< 0.2
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	< 1
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.1
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	< 10.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.3
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use


Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier

Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



Certificate of Analysis

1.00132.0000 Barbituric acid for analysis EMSURE®
Batch N020065932

	Spec. Values		Batch Values	
Assay (acidimetric)	≥ 99	%	99.6	%
Identity (IR-spectrum)	passes test		passes test	
Chloride (Cl)	≤ 40	ppm	≤ 40	ppm
Heavy metals (as Pb)	≤ 50	ppm	≤ 50	ppm
Fe (Iron)	≤ 10	ppm	≤ 10	ppm
Sulfated ash	≤ 0.1	%	≤ 0.1	%
Loss on Drying (105 °C)	≤ 0.1	%	≤ 0.1	%
Suitability as reagent (for cyanide determination)	passes test		passes test	

Date of release (DD.MM.YYYY) 17.04.2020
Minimum shelf life (DD.MM.YYYY) 30.04.2025

Ioannis Chartomatsidis
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.

Sodium Phosphate, Monobasic, Monohydrate,
Crystal
BAKER ANALYZED® A.C.S. Reagent

(sodium dihydrogen phosphate, monohydrate)



Material No.: 3818-05
Batch No.: 0000225799
Manufactured Date: 2018/12/05
Retest Date: 2025/12/03
Revision No: 1

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay ($\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$)	98.0 – 102.0 %	99.5
pH of 5% Solution at 25°C	4.1 – 4.5	4.3
Insoluble Matter	≤ 0.01 %	< 0.01
Chloride (Cl)	≤ 5 ppm	< 5
ACS – Sulfate (SO_4)	≤ 0.003 %	< 0.003
Calcium (Ca)	≤ 0.005 %	< 0.005
Potassium (K)	≤ 0.01 %	< 0.01
Heavy Metals (as Pb)	≤ 0.001 %	< 0.001
Trace Impurities – Iron (Fe)	≤ 0.001 %	< 0.001

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: IN
Packaging Site: Paris Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality


For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Certificate of Analysis

Product Name:

Zinc acetate dihydrate - ACS reagent, $\geq 98\%$

Product Number: 383058
Batch Number: MKCQ9159
Brand: SIGALD
CAS Number: 5970-45-6
MDL Number: MFCD00066961
Formula: $C_4H_6O_4Zn \cdot 2H_2O$
Formula Weight: 219.51 g/mol
Quality Release Date: 06 JAN 2022



W2926
Open 7/5/22
received
on 7/5/22

Test	Specification	Result
Appearance (Color)	White	White
Appearance (Form)	Powder or Crystal or Chunk(s)	Powder
Infrared Spectrum	Conforms to Structure	Conforms
Insoluble Matter	$\leq 0.005 \%$	0.003 %
Calcium (Ca)	$\leq 0.005 \%$	0.003 %
Chloride (Cl)	$\leq 5 \text{ ppm}$	$< 5 \text{ ppm}$
Iron (Fe)	$\leq 5 \text{ ppm}$	$< 5 \text{ ppm}$
Potassium (K)	$\leq 0.01 \%$	0.00 %
Magnesium (Mg)	$\leq 0.005 \%$	0.003 %
Sodium (Na)	$\leq 0.05 \%$	0.03 %
Lead (Pb)	$\leq 0.002 \%$	$< 0.001 \%$
pH	6.0 - 7.0	6.1
Sulfate (SO ₄)	$\leq 0.005 \%$	$< 0.005 \%$
Complexometric EDTA	98.0 - 101.0 %	100.3 %
Meets ACS Requirements	Meets Requirements	Meets Requirements


Larry Coers, Director
Quality Control
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



**RICCA CHEMICAL COMPANY®**

1490 Lammers Pike

Batesville, IN 47006

<http://www.riccachemical.com>

1-888-GO-RICCA

customerservice@riccachemical.com

Certificate of Analysis

W3093
004121
04/03/2024
16

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)**Lot Number:** 4401F99**Product Number:** 1551**Manufacture Date:** JAN 08, 2024**Expiration Date:** DEC 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	

Test	Specification	Result
Appearance	Yellow liquid	Passed

*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.004	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-1	4 L natural poly	24 months
1551-1CT	4 L Cubitainer®	24 months
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (01/08/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Certificate of Analysis

Sodium Thiosulfate, 0.0250 Normal (N/40)

Lot Number: 4403S13

Product Number: 7900

Manufacture Date: MAR 29, 2024

Expiration Date: SEP 2025

This product is specially formulated to increase its stability. A preservative is added to prevent bacterial contamination. However, all Sodium Thiosulfate solutions are subject to slow chemical deterioration and should be restandardized periodically.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Thiosulfate Pentahydrate	10102-17-7	ACS
Organic Preservative	Proprietary	
Sodium Carbonate	497-19-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Iodate/Starch)	0.02499-0.02501 N at 20°C	0.02501 N at 20°C	136

Specification	Reference
Standard Sodium Thiosulfate Solution, 0.0250 N	APHA (4500-S2- F)
Standard Sodium Thiosulfate Titrant	APHA (4500-O D)
Standard Sodium Thiosulfate Titrant	APHA (4500-O E)
Standard Sodium Thiosulfate Titrant	APHA (4500-O F)
Standard Sodium Thiosulfate Titrant, 0.025 N	APHA (4500-CI B)
Standard Sodium Thiosulfate Titrant	APHA (4500-O C)
Standard Sodium Thiosulfate Titrant, 0.025 M	APHA (5530 C)
Standard Sodium Thiosulfate Solution (0.025 N)	EPA (SW-846) (9031)
Standard Sodium Thiosulfate solution (0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7900-1	4 L natural poly	18 months
7900-16	500 mL natural poly	18 months
7900-1CT	4 L Cubitainer®	18 months
7900-32	1 L natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Certificate of Analysis



Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

Spec Set: 0583ACS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

Certificate of Analysis

Iodine (Iodine-Iodide), 0.0250 Normal (N/40), 1 mL = 0.4008 mg S²⁻**Lot Number:** 2405D89**Product Number:** 3975**Manufacture Date:** MAY 10, 2024**Expiration Date:** MAY 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Iodide	7681-11-0	ACS
Iodine	7553-56-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Dark brown liquid	Passed	
Assay (vs. Sodium Thiosulfate/Starch)	0.02498-0.02502 N at 20°C	0.02502 N at 20°C	136

Specification	Reference
Standard Iodine Solution, 0.0250 N	APHA (4500-S2- F)
Iodine Solution (approximately 0.025 N)	EPA (SW-846) (9031)
Standard Iodine Solution, 0.0250 N	EPA (376.1)
Iodine Solution (approximately 0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3975-1	4 L amber glass	12 months
3975-16	500 mL amber glass	12 months
3975-32	1 L amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jose Pena (05/10/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Part of TCP Analytical Group

Jackson's Pointe Commerce Park- Building 1000
1010 Jackson's Pointe Court, Zelienople, PA 16063

Certificate of Analysis

Cyanide Standard 1000 ppm (1ml = 1mg CN)

Product Code: **LC13545**

Manufacture Date: August 01, 2024

Lot Number: **44080060**

Expiration Date: January 30, 2025

Test	Specification	Result
Appearance (clarity)	clear solution	clear solution
Appearance (color)	colorless	colorless
Concentration (CN)	0.990 - 1.010mg/mL	1.008mg/mL
Concentration (CN)	990 - 1,010ppm	1,008ppm
Traceable to NIST SRM	Report	999b

Intended Use - Product is intended for use in manufacturing procedures and laboratory procedures and protocols.

Storage Information - Unless noted on the product label, store the product under normal lab conditions in its tightly closed, original container. Do not pipet directly from the container or return unused portions to the container.

Instructions for Handling and Use - Please refer to the associated product label and Safety Data Sheet (SDS) for information regarding safety and handling of this product.

Preparation - All products are manufactured and tested according to established, documented procedures and methodology. Production documentation records manufacturing data, raw material traceability and testing history on a per lot basis. Balances, thermometers, and glassware are calibrated before first use and on a regular schedule with references traceable to NIST standards.

The suffix of the product code may differ from what is on your product label. The suffix will designate the size and be associated with a numeric digit(s). Visit LabChem.com for more information

Suffix	1	2	3/3S/36/36S	4/4C	5	6	7	8	9	20	44	200	246	486
Size	500mL or g	1L or 1kg	2.5L/2.5L Coated/6x2.5L/6x2.5L Coated	4L	20L	10L	125mL	25g	100g	20x20mL	4x4L	200L	24x6mL	48x6mL

Michael Monteleone

Michael Monteleone
Chemistry Supervisor - Quality Control

ISO9001:2015 Registration #0306-01

2024080113:32:16bsturges-0-0

W3139 Received on 9/9/24 by IZ

Product No.: A12044
Product: Chloramine-T trihydrate, 98%
Lot No.: 10239484

Appearance:	White powder
Melting Point:	166°C(dec)
Assay (Iodometric titration):	100.5%
Identification (FTIR):	Conforms

Order our products online thermofisher.com/chemicals

This document has been electronically generated and does not require a signature.

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.



Certificate of Analysis

Starch Indicator, 0.5% (w/v), Mercury Free, for Iodometric Titrations

Lot Number: 4408P62

Product Number: 8000

Manufacture Date: AUG 28, 2024

Expiration Date: AUG 2026

This product is Mercury-free.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Starch, soluble	9005-84-9	ACS
Salicylic Acid	69-72-7	ACS

Test	Specification	Result
Appearance	White translucent liquid	Passed
Suitability for Use	Colorless (Iodine absent) - Blue (Iodine present)	Passed

Specification	Reference
Starch Solution	APHA (4500-S2- F)
Starch Indicator Solution	APHA (4500-CI B)
Starch Indicator	APHA (4500-SO32- B)
Starch indicator solution	APHA (2350 B)
Starch indicator solution	APHA (2350 E)
Starch Solution	APHA (510 B)
Starch Solution	APHA (5530 C)
Starch Indicator	APHA (4500-CI C)
Starch Indicator	EPA (345.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8000-1	4 L natural poly	24 months
8000-16	500 mL natural poly	24 months
8000-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

A handwritten signature in blue ink that reads "Paul Brandon". The signature is fluid and cursive, with the first name "Paul" and last name "Brandon" clearly distinguishable.

Paul Brandon (08/28/2024)
Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Certificate of Analysis

Cyanide Standard, 1000 ppm CN⁻

Lot Number: 1411J58**Product Number:** 2543**Manufacture Date:** NOV 22, 2024**Expiration Date:** MAY 2025

This standard is prepared using accurate volumetric techniques from material that has been assayed against Silver Nitrate solution certified traceable to NIST Standard Reference Material 999. The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is the combined uncertainty based on the stability of the assayed Potassium Cyanide, and the uncertainty in the mass and volume measurements.

Use 0.16% (w/v) (0.04 N) Sodium Hydroxide or 0.225 % (w/v) (0.04 N) Potassium Hydroxide to make dilutions of this standard. Restandardize weekly if extreme accuracy is required.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Cyanide	151-50-8	ACS
Sodium Hydroxide	1310-73-2	Reagent


Test	Specification	Result
Appearance	Colorless liquid	Passed
Cyanide (CN ⁻)	995-1005 ppm	1000 ppm

Specification	Reference
Stock Standard Cyanide Solution	APHA (4500-CN- F)
Stock Cyanide Solution	APHA (4500-CN- E)
Stock Cyanide Solution	APHA (4500-CN- K)
Stock Cyanide Solution	APHA (4500-CN- H)
Cyanide Reference Solution (1000 mg/L)	EPA (SW-846) (7.3.3.2)
Cyanide Calibration Stock Solution (1,000 mg/L CN ⁻)	EPA (SW-846) (9213)
Stock Cyanide Solution	EPA (335.3)
Stock Cyanide Solution	EPA (335.2)
Cyanide Solution Stock	ASTM (D 4282)
Simple Cyanide Solution, Stock (1.0 g/L CN ⁻)	ASTM (D 4374)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2543-16	500 mL amber poly	6 months
2543-32	1 L amber poly	6 months
2543-4	120 mL amber poly	6 months

Recommended Storage: 2°C - 8°C (36°F - 46°F)

A handwritten signature in black ink, appearing to read 'L. Briceno', with a horizontal line underneath.

Luis Briceno (11/22/2024)
Operations Supervisor

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Certificate of Analysis

Buffer, Reference Standard, pH 2.00 ± 0.01 at 25°C**Lot Number:** 2411E26**Product Number:** 1493**Manufacture Date:** NOV 11, 2024**Expiration Date:** OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	10	15	20	25	30	35	40	45	50
pH	1.93	1.98	1.98	2.00	2.01	2.03	2.03	2.04	2.04

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result
Appearance	Colorless liquid	Passed

*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	1.994	0.02	185i, 186-I-g, 186-II-g

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1493-1	4 L natural poly	24 months
1493-16	500 mL natural poly	24 months
1493-1CT	4 L Cubitainer®	24 months
1493-2.5	10 L Cubitainer®	24 months
1493-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jose Pena (11/11/2024)
Operations Manager

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

**RICCA CHEMICAL COMPANY®**

1841 Broad Street
Pocomoke City, MD 21851
<http://www.riccachemical.com>
1-888-GO-RICCA
customerservice@riccachemical.com

Certificate of Analysis

021758 58

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 2411A93

Product Number: 1501

Manufacture Date: NOV 04, 2024

Expiration Date: OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.
The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	4.00	4.00	4.00	4.00	4.00	4.00	4.01	4.02	4.03	4.04	4.06

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Acid Phthalate	877-24-7	Buffer
Preservative	Proprietary	Commercial
Red Dye	Proprietary	Purified

Test	Specification	Result
Appearance	Red liquid	Passed

*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	4.008	0.02	185i, 186-I-g, 186-II-g

Specification	Reference
Commercial Buffer Solutions	
Buffer B	ASTM (D 1293 B)
Buffer B	ASTM (D 5464)
Buffer B	ASTM (D 5128)

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1501-16	500 mL natural poly	24 months
1501-2.5	10 L Cubitainer®	24 months
1501-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



RICCA CHEMICAL COMPANY®

1841 Broad Street
Pocomoke City, MD 21851
<http://www.riccachemical.com>
1-888-GO-RICCA
customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 10.00 ± 0.01 at 25°C (Color Coded Blue)

Lot Number: 2410F80

Product Number: 1601

Manufacture Date: OCT 09, 2024

Expiration Date: MAR 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.
The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	50
pH	10.31	10.23	10.17	10.11	10.05	10.00	9.95	9.91	9.87	9.81

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Carbonate	497-19-8	ACS
Sodium Bicarbonate	144-55-8	ACS
Sodium Hydroxide	1310-73-2	Reagent
Preservative	Proprietary	
Blue Dye	Proprietary	

Test	Specification	Result
Appearance	Blue liquid	Passed

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	10.009	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	
Buffer C	ASTM (D 1293 B)
Buffer C	ASTM (D 5464)
	ASTM (D 5128)

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1601-1	4 L natural poly	18 months
1601-16	500 mL natural poly	18 months
1601-1CT	4 L Cubitainer®	18 months
1601-2.5	10 L Cubitainer®	18 months
1601-32	1 L natural poly	18 months
1601-5	20 L Cubitainer®	18 months

Version: 1.3

Lot Number: 2410F80

Product Number: 1601

Page 1 of 2



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/3/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:05
In Date: 04/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 04/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135267

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1694-01	TP-18	1	1.13	10.64	11.77	10.5	88.1	
Q1694-02	TP-18-EPH	2	1.13	10.64	11.77	9.5	78.7	
Q1694-03	TP-18-VOC	3	1.15	10.40	11.55	10.11	86.2	
Q1695-01	TT-2	4	1.14	10.33	11.47	10.21	87.8	
Q1695-02	TT-2-EPH	5	1.16	10.08	11.24	9.91	86.8	
Q1695-03	TT-2-VOC	6	1.19	10.43	11.62	10.37	88.0	
Q1695-05	TT-3	7	1.14	10.49	11.63	10.72	91.3	
Q1695-06	TT-3-EPH	8	1.19	10.73	11.92	11.01	91.5	
Q1695-07	TT-3-VOC	9	1.16	10.29	11.45	10.47	90.5	
Q1698-01	B-9	10	1.14	10.39	11.53	10.8	93.0	
Q1698-03	B-9	11	1.15	10.39	11.54	10.77	92.6	
Q1698-04	B-9-TPH-2	12	1.19	10.10	11.29	10.51	92.3	
Q1698-05	B-9-TPH-3	13	1.12	10.06	11.18	10.34	91.7	
Q1698-06	B-9-TPH-4	14	1.15	10.34	11.49	10.66	92.0	
Q1698-07	B-10	15	1.14	10.59	11.73	11.03	93.4	
Q1698-09	B-10	16	1.12	10.30	11.42	10.71	93.1	
Q1698-10	B-10-TPH-2	17	1.13	10.71	11.84	11.18	93.8	
Q1698-11	B-10-TPH-3	18	1.14	10.29	11.43	10.73	93.2	
Q1698-12	B-10-TPH-4	19	1.18	10.37	11.55	10.77	92.5	
Q1699-01	34882	20	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1701-01	4225-A	21	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1701-02	4225-B	22	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1701-03	32825-A	23	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1701-04	32825-B	24	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1705-01	ETGI-327	25	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
Q1705-02	ETGI-275	26	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
Q1706-01	PL-01-040225	27	1.13	10.28	11.41	10.37	89.9	
Q1706-02	PL-01-040225-E2	28	1.14	10.39	11.53	10.46	89.7	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/3/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:05
In Date: 04/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 04/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135267

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1706-03	PL-02-040225	29	1.16	10.61	11.77	10.63	89.3	
Q1706-04	PL-02-040225-E2	30	1.13	10.37	11.5	10.32	88.6	
Q1708-01	QQ3	31	1.00	1.00	2.00	2.00	100.0	
Q1709-01	4125-A	32	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1709-02	4125-B	33	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1709-03	4125-C	34	1.00	1.00	2.00	2.00	100.0	wipe sample

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

U135267

WorkList Name : %1-040225

WorkList ID : 188677

Department : Wet-Chemistry

Date : 04-02-2025 09:33:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1708-01	QQ3	Solid	Percent Solids	Cool 4 deg C	ALLI03		04/02/2025	Chemtech -SO
Q1698-01	B-9	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-03	B-9	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-04	B-9-TPH-2	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-05	B-9-TPH-3	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-06	B-9-TPH-4	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-07	B-10	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-09	B-10	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-10	B-10-TPH-2	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-11	B-10-TPH-3	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1698-12	B-10-TPH-4	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1694-01	TP-18	Solid	Percent Solids	Cool 4 deg C	TULL02	I41	04/02/2025	Chemtech -SO
Q1694-02	TP-18-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/01/2025	Chemtech -SO
Q1694-03	TP-18-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/01/2025	Chemtech -SO
Q1695-01	TT-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/01/2025	Chemtech -SO
Q1695-02	TT-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/02/2025	Chemtech -SO
Q1695-03	TT-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/02/2025	Chemtech -SO
Q1709-03	4125-C	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/02/2025	Chemtech -SO
Q1701-03	32825-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/02/2025	Chemtech -SO
Q1701-04	32825-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	04/02/2025	Chemtech -SO
Q1705-01	ETGI-327	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	04/02/2025	Chemtech -SO
					PSEG03	L12	04/02/2025	Chemtech -SO

Date/Time 64102125 15:30

Raw Sample Received by: TH (WOC)

Raw Sample Relinquished by: JD (CSM)

Date/Time 04102125 17:10

Raw Sample Received by: JD (CSM)

Raw Sample Relinquished by: J (WOC)

WORKLIST(Hardcopy Internal Chain)

135267

WorkList Name : %1-040225

WorkList ID : 188677

Department : Wet-Chemistry

Date : 04-02-2025 09:33:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1705-02	ETGI-275	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/02/2025	Chemtech -SO
Q1709-01	4125-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/02/2025	Chemtech -SO
Q1709-02	4125-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/02/2025	Chemtech -SO
Q1695-05	TT-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/02/2025	Chemtech -SO
Q1695-06	TT-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/02/2025	Chemtech -SO
Q1695-07	TT-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/02/2025	Chemtech -SO
Q1699-01	34882	Solid	Percent Solids	Cool 4 deg C	PSEG03	I41	04/02/2025	Chemtech -SO
Q1701-01	4225-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	04/02/2025	Chemtech -SO
Q1701-02	4225-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	04/02/2025	Chemtech -SO
Q1706-01	PL-01-040225	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	04/02/2025	Chemtech -SO
Q1706-02	PL-01-040225-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	I31	04/02/2025	Chemtech -SO
Q1706-03	PL-02-040225	Solid	Percent Solids	Cool 4 deg C	PSEG05	I31	04/02/2025	Chemtech -SO
Q1706-04	PL-02-040225-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	I31	04/02/2025	Chemtech -SO

Date/Time 04/02/25 15:30

Raw Sample Received by: SR

Raw Sample Relinquished by: JD (SM)

Date/Time 04/02/25 17:10

Raw Sample Received by: JD (SM)

Raw Sample Relinquished by: JD (SM)



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: Tully Construction Co. Inc

ADDRESS: Rockway Freeway & Beach Channel

CITY: Queens STATE: NY ZIP: 11416

ATTENTION: Dean Devoe

PHONE:

FAX:

PROJECT NAME:

PROJECT NO.: LOCATION:

PROJECT MANAGER:

e-mail:

PHONE:

FAX:

BILL TO:

PO#:

ADDRESS:

CITY

STATE:

ZIP:

ATTENTION:

PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*

HARDCOPY (DATA PACKAGE): DAYS*

EDD: DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)

☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP

☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B

+ Raw Data ☐ Other

☐ EDD FORMAT

1: Vol's 2: SWA 3: PCB 4: TAP Extraction 5: Corrosivity 6: Ignitability 7: Reactivity Cyn 8: Reactive Sulfate 9: Mercury Metals TPH

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		E/F	E	E	E	E	E	E	E	E	← Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-OTHER	
1.	B-9	SOL	X		040225	0810	8		X	X	X	X	X	X	X	X		
2.	B-9			X		0815	4	X									0.0 ppm	
3.	B-9-TPH-2		X			0817	1									X		
4.	B-9-TPH-3		X			0819	1									X		
5.	B-9-TPH-4		X			0822	1									X		
6.	B-10		X			0835	8		X	X	X	X	X	X	X	X		
7.	B-10			X		0839	4	X									0.0 ppm	
8.	B-10-TPH-2		X			0842	1									X		
9.	B-10-TPH-3		X			0845	1									X		
10.	B-10-TPH-4		X			0848	1									X		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME: 0930	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP 2.2 °C
1. <i>[Signature]</i>	04-02-25	1. <i>[Signature]</i>	Comments: Collected (2) 8:1 Composite sample + (4) 5:1 Composite samples for sampling protocol TPH.
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
2. <i>[Signature]</i>		2. <i>[Signature]</i>	
RELINQUISHED BY SAMPLER:	DATE/TIME: 1200	RECEIVED BY:	PTD Calibration 04-02-25
3. <i>[Signature]</i>	04-02-25	3. <i>[Signature]</i>	Client: ☐ Hand Delivered ☐ Other
			Shipment Complete ☒ YES ☐ NO

04-02-25



Legend

 Boring Location

 Approximate Property Boundary

0 95 190 380 570 Feet

MATRIX^{NEW}**WORLD**
Engineering Progress

Matrix New World Engineering, Land Surveying
and Landscape Architecture, P.C.
333 West 39th Street, Suite 202
New York, New York 10018
Tel: 973-240-1800
Fax: 973-240-1818
WBE / DBE / SBE
www.matrixnewworld.com
STATE OF NEW YORK CERTIFICATE OF AUTHORIZATION No. 17-082661

AS DRILLED BORING LOCATION MAP

SUBSTATION 117
ROCKAWAY FREEWAY & BEACH CHANNEL DRIVE
BLOCK 16186, LOT 100
ROCKAWAY PARK, QUEENS, NEW YORK

JUNE 2019

FIGURE 4

Source: Esri, DigitalGlobe, GeoEye, Earthstar Geographics, CNES/Airbus DS, USDA, USGS, AeroGRID, IGN, and the GIS User Community



Environmental Laboratory

www.chemtech.net | EMAIL: PM@chemtech.net

Project Name: _____ Chemtech Order ID: _____
Service Order #: _____ Sampler Name: Jeremy M
Work Order #: _____ Client Project Coordinator & Phone: Dean Devoe
Labor WBS #: _____ Page #: 1 of 1
Facility/Site: Substation 117 Date: 04-02-25
Site Address: Lockway freeway & Arrive Time: 0800
Beach Channel Dr, Queens NY Depart Time: 0930

Waste Stream (circle one): drum / roll-off / soil pile / in-situ / linear construction / frac-tank

Sample Matrices (circle all that apply): Water / Solid / NAPL / Concrete / Wipe

Collection Depths: N/A Dimensions/CY: N/A

Temp (range): 2-2 °C PID Readings (range): 0.0 PPM Odor: Y ☒ Color: Y ☒ N

Sample Description: Dark Brown Soil, Sand, Gravel.

Field Observations: (2) location (B-9 & B10)

Grid/Area Composite Map:

QA Control # A3041134

See

Attached

MAP *

Sampler Signature: pm

Supervisor Review/Date: _____

Client Signature: _____

Date/Time Arrived at Lab: _____

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1698	TULL02	Order Date : 4/2/2025 12:14:00 PM	Project Mgr :
Client Name : Tully Construction Co., Inc.		Project Name : MTA Rockaway Park	Report Type : Level 1
Client Contact : Dean Devoe		Receive DateTime : 4/2/2025 12:00:00 PM	EDD Type : Excel NY 375
Invoice Name : Tully Construction Co., Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Dean Devoe			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1698-03	B-9	Solid	04/02/2025	08:15					
					VOC-TCLVOA-10		8260D	5 Bus. Days	
Q1698-09	B-10	Solid	04/02/2025	08:39					
					VOC-TCLVOA-10		8260D	5 Bus. Days	

Relinquished By :

21

Date / Time :

04-02-25 1330

Received By :

Date / Time :

pl
04.02.25 13:30

Storage Area : VOA Refridgerator Room