



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Tully Construction Co., Inc.				SDG No.:	Q1698			
Contract:	TULL02	Lab Code:	CHEM		Case No.:	Q1698	SAS No.:	Q1698	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB79	Mercury	0.20	+/-0.20	U	0.20	CV	04/07/2025	13:32	LB135320

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB39	Mercury	0.20	+/-0.20	U	0.20	CV	04/07/2025	13:36	LB135320
CCB40	Mercury	0.20	+/-0.20	U	0.20	CV	04/07/2025	14:04	LB135320
CCB41	Mercury	0.20	+/-0.20	U	0.20	CV	04/07/2025	14:37	LB135320

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	04/07/2025	13:05	LB135333
	Antimony	50.0	+/-50.0	U	50.0	P	04/07/2025	13:05	LB135333
	Arsenic	20.0	+/-20.0	U	20.0	P	04/07/2025	13:05	LB135333
	Barium	100	+/-100	U	100	P	04/07/2025	13:05	LB135333
	Beryllium	6.00	+/-6.00	U	6.00	P	04/07/2025	13:05	LB135333
	Cadmium	6.00	+/-6.00	U	6.00	P	04/07/2025	13:05	LB135333
	Calcium	2000	+/-2000	U	2000	P	04/07/2025	13:05	LB135333
	Chromium	10.0	+/-10.0	U	10.0	P	04/07/2025	13:05	LB135333
	Cobalt	30.0	+/-30.0	U	30.0	P	04/07/2025	13:05	LB135333
	Copper	20.0	+/-20.0	U	20.0	P	04/07/2025	13:05	LB135333
	Iron	100	+/-100	U	100	P	04/07/2025	13:05	LB135333
	Lead	12.0	+/-12.0	U	12.0	P	04/07/2025	13:05	LB135333
	Magnesium	2000	+/-2000	U	2000	P	04/07/2025	13:05	LB135333
	Manganese	20.0	+/-20.0	U	20.0	P	04/07/2025	13:05	LB135333
	Nickel	40.0	+/-40.0	U	40.0	P	04/07/2025	13:05	LB135333
	Potassium	2000	+/-2000	U	2000	P	04/07/2025	13:05	LB135333
	Selenium	20.0	+/-20.0	U	20.0	P	04/07/2025	13:05	LB135333
	Silver	10.0	+/-10.0	U	10.0	P	04/07/2025	13:05	LB135333
	Sodium	2000	+/-2000	U	2000	P	04/07/2025	13:05	LB135333
	Thallium	40.0	+/-40.0	U	40.0	P	04/07/2025	13:05	LB135333
	Vanadium	40.0	+/-40.0	U	40.0	P	04/07/2025	13:05	LB135333
	Zinc	40.0	+/-40.0	U	40.0	P	04/07/2025	13:05	LB135333

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	04/07/2025	13:50	LB135333
	Antimony	50.0	+/-50.0	U	50.0	P	04/07/2025	13:50	LB135333
	Arsenic	20.0	+/-20.0	U	20.0	P	04/07/2025	13:50	LB135333
	Barium	100	+/-100	U	100	P	04/07/2025	13:50	LB135333
	Beryllium	6.00	+/-6.00	U	6.00	P	04/07/2025	13:50	LB135333
	Cadmium	6.00	+/-6.00	U	6.00	P	04/07/2025	13:50	LB135333
	Calcium	2000	+/-2000	U	2000	P	04/07/2025	13:50	LB135333
	Chromium	10.0	+/-10.0	U	10.0	P	04/07/2025	13:50	LB135333
	Cobalt	30.0	+/-30.0	U	30.0	P	04/07/2025	13:50	LB135333
	Copper	20.0	+/-20.0	U	20.0	P	04/07/2025	13:50	LB135333
	Iron	100	+/-100	U	100	P	04/07/2025	13:50	LB135333
	Lead	12.0	+/-12.0	U	12.0	P	04/07/2025	13:50	LB135333
	Magnesium	2000	+/-2000	U	2000	P	04/07/2025	13:50	LB135333
	Manganese	20.0	+/-20.0	U	20.0	P	04/07/2025	13:50	LB135333
	Nickel	40.0	+/-40.0	U	40.0	P	04/07/2025	13:50	LB135333
	Potassium	2000	+/-2000	U	2000	P	04/07/2025	13:50	LB135333
	Selenium	20.0	+/-20.0	U	20.0	P	04/07/2025	13:50	LB135333
	Silver	10.0	+/-10.0	U	10.0	P	04/07/2025	13:50	LB135333
	Sodium	2000	+/-2000	U	2000	P	04/07/2025	13:50	LB135333
	Thallium	40.0	+/-40.0	U	40.0	P	04/07/2025	13:50	LB135333
	Vanadium	40.0	+/-40.0	U	40.0	P	04/07/2025	13:50	LB135333
	Zinc	40.0	+/-40.0	U	40.0	P	04/07/2025	13:50	LB135333
CCB02	Aluminum	100	+/-100	U	100	P	04/07/2025	15:12	LB135333
	Antimony	50.0	+/-50.0	U	50.0	P	04/07/2025	15:12	LB135333
	Arsenic	20.0	+/-20.0	U	20.0	P	04/07/2025	15:12	LB135333
	Barium	100	+/-100	U	100	P	04/07/2025	15:12	LB135333
	Beryllium	6.00	+/-6.00	U	6.00	P	04/07/2025	15:12	LB135333
	Cadmium	6.00	+/-6.00	U	6.00	P	04/07/2025	15:12	LB135333
	Calcium	2000	+/-2000	U	2000	P	04/07/2025	15:12	LB135333
	Chromium	10.0	+/-10.0	U	10.0	P	04/07/2025	15:12	LB135333
	Cobalt	30.0	+/-30.0	U	30.0	P	04/07/2025	15:12	LB135333
	Copper	20.0	+/-20.0	U	20.0	P	04/07/2025	15:12	LB135333
	Iron	100	+/-100	U	100	P	04/07/2025	15:12	LB135333
	Lead	12.0	+/-12.0	U	12.0	P	04/07/2025	15:12	LB135333
	Magnesium	2000	+/-2000	U	2000	P	04/07/2025	15:12	LB135333
	Manganese	20.0	+/-20.0	U	20.0	P	04/07/2025	15:12	LB135333
	Nickel	40.0	+/-40.0	U	40.0	P	04/07/2025	15:12	LB135333
	Potassium	2000	+/-2000	U	2000	P	04/07/2025	15:12	LB135333
	Selenium	20.0	+/-20.0	U	20.0	P	04/07/2025	15:12	LB135333

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	04/07/2025	15:12	LB135333
	Sodium	2000	+/-2000	U	2000	P	04/07/2025	15:12	LB135333
	Thallium	40.0	+/-40.0	U	40.0	P	04/07/2025	15:12	LB135333
	Vanadium	40.0	+/-40.0	U	40.0	P	04/07/2025	15:12	LB135333
	Zinc	40.0	+/-40.0	U	40.0	P	04/07/2025	15:12	LB135333
CCB03	Aluminum	100	+/-100	U	100	P	04/07/2025	15:45	LB135333
	Antimony	50.0	+/-50.0	U	50.0	P	04/07/2025	15:45	LB135333
	Arsenic	20.0	+/-20.0	U	20.0	P	04/07/2025	15:45	LB135333
	Barium	100	+/-100	U	100	P	04/07/2025	15:45	LB135333
	Beryllium	6.00	+/-6.00	U	6.00	P	04/07/2025	15:45	LB135333
	Cadmium	6.00	+/-6.00	U	6.00	P	04/07/2025	15:45	LB135333
	Calcium	2000	+/-2000	U	2000	P	04/07/2025	15:45	LB135333
	Chromium	10.0	+/-10.0	U	10.0	P	04/07/2025	15:45	LB135333
	Cobalt	30.0	+/-30.0	U	30.0	P	04/07/2025	15:45	LB135333
	Copper	20.0	+/-20.0	U	20.0	P	04/07/2025	15:45	LB135333
	Iron	100	+/-100	U	100	P	04/07/2025	15:45	LB135333
	Lead	12.0	+/-12.0	U	12.0	P	04/07/2025	15:45	LB135333
	Magnesium	2000	+/-2000	U	2000	P	04/07/2025	15:45	LB135333
	Manganese	20.0	+/-20.0	U	20.0	P	04/07/2025	15:45	LB135333
	Nickel	40.0	+/-40.0	U	40.0	P	04/07/2025	15:45	LB135333
	Potassium	2000	+/-2000	U	2000	P	04/07/2025	15:45	LB135333
	Selenium	20.0	+/-20.0	U	20.0	P	04/07/2025	15:45	LB135333
	Silver	10.0	+/-10.0	U	10.0	P	04/07/2025	15:45	LB135333
	Sodium	2000	+/-2000	U	2000	P	04/07/2025	15:45	LB135333
	Thallium	40.0	+/-40.0	U	40.0	P	04/07/2025	15:45	LB135333
	Vanadium	40.0	+/-40.0	U	40.0	P	04/07/2025	15:45	LB135333
	Zinc	40.0	+/-40.0	U	40.0	P	04/07/2025	15:45	LB135333
CCB04	Aluminum	100	+/-100	U	100	P	04/07/2025	17:04	LB135333
	Antimony	50.0	+/-50.0	U	50.0	P	04/07/2025	17:04	LB135333
	Arsenic	20.0	+/-20.0	U	20.0	P	04/07/2025	17:04	LB135333
	Barium	100	+/-100	U	100	P	04/07/2025	17:04	LB135333
	Beryllium	6.00	+/-6.00	U	6.00	P	04/07/2025	17:04	LB135333
	Cadmium	6.00	+/-6.00	U	6.00	P	04/07/2025	17:04	LB135333
	Calcium	2000	+/-2000	U	2000	P	04/07/2025	17:04	LB135333
	Chromium	10.0	+/-10.0	U	10.0	P	04/07/2025	17:04	LB135333
	Cobalt	30.0	+/-30.0	U	30.0	P	04/07/2025	17:04	LB135333
	Copper	20.0	+/-20.0	U	20.0	P	04/07/2025	17:04	LB135333
	Iron	100	+/-100	U	100	P	04/07/2025	17:04	LB135333
	Lead	12.0	+/-12.0	U	12.0	P	04/07/2025	17:04	LB135333

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	04/07/2025	17:04	LB135333
	Manganese	20.0	+/-20.0	U	20.0	P	04/07/2025	17:04	LB135333
	Nickel	40.0	+/-40.0	U	40.0	P	04/07/2025	17:04	LB135333
	Potassium	2000	+/-2000	U	2000	P	04/07/2025	17:04	LB135333
	Selenium	20.0	+/-20.0	U	20.0	P	04/07/2025	17:04	LB135333
	Silver	10.0	+/-10.0	U	10.0	P	04/07/2025	17:04	LB135333
	Sodium	2000	+/-2000	U	2000	P	04/07/2025	17:04	LB135333
	Thallium	40.0	+/-40.0	U	40.0	P	04/07/2025	17:04	LB135333
	Vanadium	40.0	+/-40.0	U	40.0	P	04/07/2025	17:04	LB135333
	Zinc	40.0	+/-40.0	U	40.0	P	04/07/2025	17:04	LB135333
CCB05	Aluminum	100	+/-100	U	100	P	04/07/2025	17:47	LB135333
	Antimony	50.0	+/-50.0	U	50.0	P	04/07/2025	17:47	LB135333
	Arsenic	20.0	+/-20.0	U	20.0	P	04/07/2025	17:47	LB135333
	Barium	100	+/-100	U	100	P	04/07/2025	17:47	LB135333
	Beryllium	6.00	+/-6.00	U	6.00	P	04/07/2025	17:47	LB135333
	Cadmium	6.00	+/-6.00	U	6.00	P	04/07/2025	17:47	LB135333
	Calcium	2000	+/-2000	U	2000	P	04/07/2025	17:47	LB135333
	Chromium	10.0	+/-10.0	U	10.0	P	04/07/2025	17:47	LB135333
	Cobalt	30.0	+/-30.0	U	30.0	P	04/07/2025	17:47	LB135333
	Copper	20.0	+/-20.0	U	20.0	P	04/07/2025	17:47	LB135333
	Iron	100	+/-100	U	100	P	04/07/2025	17:47	LB135333
	Lead	12.0	+/-12.0	U	12.0	P	04/07/2025	17:47	LB135333
	Magnesium	2000	+/-2000	U	2000	P	04/07/2025	17:47	LB135333
	Manganese	20.0	+/-20.0	U	20.0	P	04/07/2025	17:47	LB135333
	Nickel	40.0	+/-40.0	U	40.0	P	04/07/2025	17:47	LB135333
	Potassium	2000	+/-2000	U	2000	P	04/07/2025	17:47	LB135333
	Selenium	20.0	+/-20.0	U	20.0	P	04/07/2025	17:47	LB135333
	Silver	10.0	+/-10.0	U	10.0	P	04/07/2025	17:47	LB135333
	Sodium	2000	+/-2000	U	2000	P	04/07/2025	17:47	LB135333
	Thallium	40.0	+/-40.0	U	40.0	P	04/07/2025	17:47	LB135333
	Vanadium	40.0	+/-40.0	U	40.0	P	04/07/2025	17:47	LB135333
	Zinc	40.0	+/-40.0	U	40.0	P	04/07/2025	17:47	LB135333
CCB06	Aluminum	100	+/-100	U	100	P	04/07/2025	18:55	LB135333
	Antimony	50.0	+/-50.0	U	50.0	P	04/07/2025	18:55	LB135333
	Arsenic	20.0	+/-20.0	U	20.0	P	04/07/2025	18:55	LB135333
	Barium	100	+/-100	U	100	P	04/07/2025	18:55	LB135333
	Beryllium	6.00	+/-6.00	U	6.00	P	04/07/2025	18:55	LB135333
	Cadmium	6.00	+/-6.00	U	6.00	P	04/07/2025	18:55	LB135333
	Calcium	2000	+/-2000	U	2000	P	04/07/2025	18:55	LB135333

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** Q1698
Contract: TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	04/07/2025	18:55	LB135333
	Cobalt	30.0	+/-30.0	U	30.0	P	04/07/2025	18:55	LB135333
	Copper	20.0	+/-20.0	U	20.0	P	04/07/2025	18:55	LB135333
	Iron	100	+/-100	U	100	P	04/07/2025	18:55	LB135333
	Lead	12.0	+/-12.0	U	12.0	P	04/07/2025	18:55	LB135333
	Magnesium	2000	+/-2000	U	2000	P	04/07/2025	18:55	LB135333
	Manganese	20.0	+/-20.0	U	20.0	P	04/07/2025	18:55	LB135333
	Nickel	40.0	+/-40.0	U	40.0	P	04/07/2025	18:55	LB135333
	Potassium	2000	+/-2000	U	2000	P	04/07/2025	18:55	LB135333
	Selenium	20.0	+/-20.0	U	20.0	P	04/07/2025	18:55	LB135333
	Silver	10.0	+/-10.0	U	10.0	P	04/07/2025	18:55	LB135333
	Sodium	2000	+/-2000	U	2000	P	04/07/2025	18:55	LB135333
	Thallium	40.0	+/-40.0	U	40.0	P	04/07/2025	18:55	LB135333
	Vanadium	40.0	+/-40.0	U	40.0	P	04/07/2025	18:55	LB135333
	Zinc	40.0	+/-40.0	U	40.0	P	04/07/2025	18:55	LB135333

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Tully Construction Co., Inc.</u>				SDG No.:	<u>Q1698</u>			
Contract:	<u>TULL02</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1698</u>	SAS No.:	<u>Q1698</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number