

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** Q1698
Contract: TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167441BS							
Aluminum	mg/Kg	92.6	85.5		92	80 - 120	P
Antimony	mg/Kg	37.0	35.9		97	80 - 120	P
Arsenic	mg/Kg	37.0	34.3		93	80 - 120	P
Barium	mg/Kg	9.3	8.50		91	80 - 120	P
Beryllium	mg/Kg	9.3	8.84		95	80 - 120	P
Cadmium	mg/Kg	9.3	8.58		92	80 - 120	P
Calcium	mg/Kg	46.3	44.0	J	95	80 - 120	P
Chromium	mg/Kg	18.5	18.2		98	80 - 120	P
Cobalt	mg/Kg	9.3	8.68		93	80 - 120	P
Copper	mg/Kg	13.9	13.8		99	80 - 120	P
Iron	mg/Kg	140	134		96	80 - 120	P
Lead	mg/Kg	46.3	42.4		92	80 - 120	P
Magnesium	mg/Kg	92.6	84.0	J	91	80 - 120	P
Manganese	mg/Kg	9.3	8.76		94	80 - 120	P
Nickel	mg/Kg	23.1	21.8		94	80 - 120	P
Potassium	mg/Kg	460	425		92	80 - 120	P
Selenium	mg/Kg	92.6	87.1		94	80 - 120	P
Silver	mg/Kg	3.5	3.42		98	80 - 120	P
Sodium	mg/Kg	140	138		99	80 - 120	P
Thallium	mg/Kg	92.6	89.7		97	80 - 120	P
Vanadium	mg/Kg	13.9	13.0		94	80 - 120	P
Zinc	mg/Kg	9.3	9.11		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167478BS Mercury	mg/Kg	0.27	0.25		92	80 - 120	CV