

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Construction Co., Inc. **SDG No.:** Q1698  
**Contract:** TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|----|------------------|------------------|---------------|
| ICV76     | Mercury | 3.73           | 4.0        | 93            | 90 - 110                  | CV | 04/04/2025       | 12:11            | LB135304      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Construction Co., Inc. **SDG No.:** Q1698  
**Contract:** TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|----|------------------|------------------|---------------|
| CCV30     | Mercury | 4.74           | 5.0        | 95            | 90 - 110                  | CV | 04/04/2025       | 12:16            | LB135304      |
| CCV31     | Mercury | 4.58           | 5.0        | 92            | 90 - 110                  | CV | 04/04/2025       | 12:46            | LB135304      |
| CCV32     | Mercury | 4.60           | 5.0        | 92            | 90 - 110                  | CV | 04/04/2025       | 13:08            | LB135304      |
| CCV33     | Mercury | 4.60           | 5.0        | 92            | 90 - 110                  | CV | 04/04/2025       | 13:20            | LB135304      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Construction Co., Inc. **SDG No.:** Q1698  
**Contract:** TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Arsenic  | 1030           | 1000       | 103           | 90 - 110                  | P | 04/04/2025       | 09:50            | LB135316      |
|           | Barium   | 505            | 520        | 97            | 90 - 110                  | P | 04/04/2025       | 09:50            | LB135316      |
|           | Cadmium  | 522            | 510        | 102           | 90 - 110                  | P | 04/04/2025       | 09:50            | LB135316      |
|           | Chromium | 551            | 520        | 106           | 90 - 110                  | P | 04/04/2025       | 09:50            | LB135316      |
|           | Lead     | 1030           | 1000       | 103           | 90 - 110                  | P | 04/04/2025       | 09:50            | LB135316      |
|           | Selenium | 1070           | 1000       | 107           | 90 - 110                  | P | 04/04/2025       | 09:50            | LB135316      |
|           | Silver   | 227            | 250        | 91            | 90 - 110                  | P | 04/04/2025       | 09:50            | LB135316      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

Client: Tully Construction Co., Inc.      SDG No.: Q1698

Contract: TULL02      Lab Code: CHEM      Case No.: Q1698      SAS No.: Q1698

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

---

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| LLICV01   | Arsenic  | 18.3           | 20.0       | 92            | 80 - 120                  | P | 04/04/2025       | 10:02            | LB135316      |
|           | Barium   | 92.0           | 100        | 92            | 80 - 120                  | P | 04/04/2025       | 10:02            | LB135316      |
|           | Cadmium  | 5.90           | 6.0        | 98            | 80 - 120                  | P | 04/04/2025       | 10:02            | LB135316      |
|           | Chromium | 10.2           | 10.0       | 102           | 80 - 120                  | P | 04/04/2025       | 10:02            | LB135316      |
|           | Lead     | 11.1           | 12.0       | 92            | 80 - 120                  | P | 04/04/2025       | 10:02            | LB135316      |
|           | Selenium | 19.5           | 20.0       | 97            | 80 - 120                  | P | 04/04/2025       | 10:02            | LB135316      |
|           | Silver   | 10.7           | 10.0       | 107           | 80 - 120                  | P | 04/04/2025       | 10:02            | LB135316      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Construction Co., Inc. **SDG No.:** Q1698  
**Contract:** TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Arsenic  | 4810           | 5000       | 96            | 90 - 110                  | P | 04/04/2025       | 10:41            | LB135316      |
|           | Barium   | 9100           | 10000      | 91            | 90 - 110                  | P | 04/04/2025       | 10:41            | LB135316      |
|           | Cadmium  | 2400           | 2500       | 96            | 90 - 110                  | P | 04/04/2025       | 10:41            | LB135316      |
|           | Chromium | 989            | 1000       | 99            | 90 - 110                  | P | 04/04/2025       | 10:41            | LB135316      |
|           | Lead     | 4820           | 5000       | 96            | 90 - 110                  | P | 04/04/2025       | 10:41            | LB135316      |
|           | Selenium | 4810           | 5000       | 96            | 90 - 110                  | P | 04/04/2025       | 10:41            | LB135316      |
|           | Silver   | 1200           | 1250       | 96            | 90 - 110                  | P | 04/04/2025       | 10:41            | LB135316      |
| CCV02     | Arsenic  | 4980           | 5000       | 100           | 90 - 110                  | P | 04/04/2025       | 11:36            | LB135316      |
|           | Barium   | 9220           | 10000      | 92            | 90 - 110                  | P | 04/04/2025       | 11:36            | LB135316      |
|           | Cadmium  | 2440           | 2500       | 98            | 90 - 110                  | P | 04/04/2025       | 11:36            | LB135316      |
|           | Chromium | 1030           | 1000       | 103           | 90 - 110                  | P | 04/04/2025       | 11:36            | LB135316      |
|           | Lead     | 4940           | 5000       | 99            | 90 - 110                  | P | 04/04/2025       | 11:36            | LB135316      |
|           | Selenium | 5020           | 5000       | 100           | 90 - 110                  | P | 04/04/2025       | 11:36            | LB135316      |
|           | Silver   | 1250           | 1250       | 100           | 90 - 110                  | P | 04/04/2025       | 11:36            | LB135316      |
| CCV03     | Arsenic  | 4940           | 5000       | 99            | 90 - 110                  | P | 04/04/2025       | 12:38            | LB135316      |
|           | Barium   | 9290           | 10000      | 93            | 90 - 110                  | P | 04/04/2025       | 12:38            | LB135316      |
|           | Cadmium  | 2450           | 2500       | 98            | 90 - 110                  | P | 04/04/2025       | 12:38            | LB135316      |
|           | Chromium | 1030           | 1000       | 103           | 90 - 110                  | P | 04/04/2025       | 12:38            | LB135316      |
|           | Lead     | 4960           | 5000       | 99            | 90 - 110                  | P | 04/04/2025       | 12:38            | LB135316      |
|           | Selenium | 4970           | 5000       | 99            | 90 - 110                  | P | 04/04/2025       | 12:38            | LB135316      |
|           | Silver   | 1250           | 1250       | 100           | 90 - 110                  | P | 04/04/2025       | 12:38            | LB135316      |