



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Tully Construction Co., Inc.</u>				SDG No.:	<u>Q1698</u>			
Contract:	<u>TULL02</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1698</u>	SAS No.:	<u>Q1698</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB76	Mercury	0.20	+/-0.20	U	0.20	CV	04/04/2025	12:13	LB135304

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB30	Mercury	0.20	+/-0.20	U	0.20	CV	04/04/2025	12:18	LB135304
CCB31	Mercury	0.20	+/-0.20	U	0.20	CV	04/04/2025	12:48	LB135304
CCB32	Mercury	0.20	+/-0.20	U	0.20	CV	04/04/2025	13:11	LB135304
CCB33	Mercury	0.20	+/-0.20	U	0.20	CV	04/04/2025	13:22	LB135304

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Construction Co., Inc.</u>		SDG No.: <u>Q1698</u>							
Contract: <u>TULL02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1698</u>	SAS No.: <u>Q1698</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/04/2025	10:06	LB135316
	Barium	100	+/-100	U	100	P	04/04/2025	10:06	LB135316
	Cadmium	6.00	+/-6.00	U	6.00	P	04/04/2025	10:06	LB135316
	Chromium	10.0	+/-10.0	U	10.0	P	04/04/2025	10:06	LB135316
	Lead	12.0	+/-12.0	U	12.0	P	04/04/2025	10:06	LB135316
	Selenium	20.0	+/-20.0	U	20.0	P	04/04/2025	10:06	LB135316
	Silver	10.0	+/-10.0	U	10.0	P	04/04/2025	10:06	LB135316

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** Q1698
Contract: TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/04/2025	10:49	LB135316
	Barium	100	+/-100	U	100	P	04/04/2025	10:49	LB135316
	Cadmium	6.00	+/-6.00	U	6.00	P	04/04/2025	10:49	LB135316
	Chromium	10.0	+/-10.0	U	10.0	P	04/04/2025	10:49	LB135316
	Lead	12.0	+/-12.0	U	12.0	P	04/04/2025	10:49	LB135316
	Selenium	20.0	+/-20.0	U	20.0	P	04/04/2025	10:49	LB135316
	Silver	10.0	+/-10.0	U	10.0	P	04/04/2025	10:49	LB135316
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	04/04/2025	11:41	LB135316
	Barium	100	+/-100	U	100	P	04/04/2025	11:41	LB135316
	Cadmium	6.00	+/-6.00	U	6.00	P	04/04/2025	11:41	LB135316
	Chromium	10.0	+/-10.0	U	10.0	P	04/04/2025	11:41	LB135316
	Lead	12.0	+/-12.0	U	12.0	P	04/04/2025	11:41	LB135316
	Selenium	20.0	+/-20.0	U	20.0	P	04/04/2025	11:41	LB135316
	Silver	10.0	+/-10.0	U	10.0	P	04/04/2025	11:41	LB135316
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	04/04/2025	12:42	LB135316
	Barium	100	+/-100	U	100	P	04/04/2025	12:42	LB135316
	Cadmium	6.00	+/-6.00	U	6.00	P	04/04/2025	12:42	LB135316
	Chromium	10.0	+/-10.0	U	10.0	P	04/04/2025	12:42	LB135316
	Lead	12.0	+/-12.0	U	12.0	P	04/04/2025	12:42	LB135316
	Selenium	20.0	+/-20.0	U	20.0	P	04/04/2025	12:42	LB135316
	Silver	10.0	+/-10.0	U	10.0	P	04/04/2025	12:42	LB135316

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------