

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc. **SDG No.:** Q1698
Contract: TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	3.57			-20	20	04/04/2025	10:22	LB135316
	Barium	1.87	6.0	31	-94	106	04/04/2025	10:22	LB135316
	Cadmium	4.41	1.0	441	-5	7	04/04/2025	10:22	LB135316
	Chromium	55.0	52.0	106	42	62	04/04/2025	10:22	LB135316
	Lead	11.3			-12	12	04/04/2025	10:22	LB135316
	Selenium	-1.37			-20	20	04/04/2025	10:22	LB135316
	Silver	1.96			-10	10	04/04/2025	10:22	LB135316
ICSAB01	Arsenic	111	104	107	88.4	120	04/04/2025	10:26	LB135316
	Barium	473	537	88	437	637	04/04/2025	10:26	LB135316
	Cadmium	1010	972	104	826	1120	04/04/2025	10:26	LB135316
	Chromium	583	542	108	460	624	04/04/2025	10:26	LB135316
	Lead	57.4	49.0	117	37	61	04/04/2025	10:26	LB135316
	Selenium	43.9	46.0	95	26	66	04/04/2025	10:26	LB135316
	Silver	195	201	97	170	232	04/04/2025	10:26	LB135316