

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1698

Contract: TULL02

Lab Code: CHEM

Case No.: Q1698

SAS No.: Q1698

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167446BS Mercury	ug/L	4.0	3.73		93	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** Q1698
Contract: TULL02 **Lab Code:** CHEM **Case No.:** Q1698 **SAS No.:** Q1698

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167447BS							
Arsenic	ug/L	4000	3830		96	80 - 120	P
Barium	ug/L	1000	885		88	80 - 120	P
Cadmium	ug/L	1000	954		95	80 - 120	P
Chromium	ug/L	2000	2050		102	80 - 120	P
Lead	ug/L	5000	4750		95	80 - 120	P
Selenium	ug/L	10000	9720		97	80 - 120	P
Silver	ug/L	380	373		98	80 - 120	P