

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 04/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP071050.D Time Analyzed: 09:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.672	5.563	5.763	477.180	500.000	-4.6
Aroclor-1016-2	5.694	5.585	5.785	478.590	500.000	-4.3
Aroclor-1016-3	5.756	5.648	5.848	478.190	500.000	-4.4
Aroclor-1016-4	5.853	5.745	5.945	510.100	500.000	2.0
Aroclor-1016-5	6.146	6.038	6.238	532.390	500.000	6.5
Aroclor-1260-1	7.266	7.156	7.356	575.670	500.000	15.1
Aroclor-1260-2	7.519	7.410	7.610	553.250	500.000	10.7
Aroclor-1260-3	7.879	7.768	7.968	528.920	500.000	5.8
Aroclor-1260-4	8.103	7.992	8.192	506.530	500.000	1.3
Aroclor-1260-5	8.424	8.311	8.511	526.120	500.000	5.2
Decachlorobiphenyl	10.243	10.125	10.325	53.060	50.000	6.1
Tetrachloro-m-xylene	4.518	4.412	4.612	49.510	50.000	-1.0

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 04/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP071050.D Time Analyzed: 09:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.902	4.803	5.003	481.520	500.000	-3.7
Aroclor-1016-2	4.920	4.822	5.022	466.920	500.000	-6.6
Aroclor-1016-3	5.097	4.999	5.199	498.680	500.000	-0.3
Aroclor-1016-4	5.139	5.041	5.241	498.280	500.000	-0.3
Aroclor-1016-5	5.353	5.255	5.455	522.950	500.000	4.6
Aroclor-1260-1	6.389	6.292	6.492	535.630	500.000	7.1
Aroclor-1260-2	6.577	6.480	6.680	535.270	500.000	7.1
Aroclor-1260-3	6.731	6.634	6.834	532.370	500.000	6.5
Aroclor-1260-4	7.202	7.105	7.305	500.620	500.000	0.1
Aroclor-1260-5	7.443	7.346	7.546	455.560	500.000	-8.9
Decachlorobiphenyl	8.856	8.759	8.959	46.200	50.000	-7.6
Tetrachloro-m-xylene	3.815	3.716	3.916	45.900	50.000	-8.2