

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 04/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP071065.D Time Analyzed: 15:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.674	5.563	5.763	491.200	500.000	-1.8
Aroclor-1016-2	5.695	5.585	5.785	458.810	500.000	-8.2
Aroclor-1016-3	5.758	5.648	5.848	489.710	500.000	-2.1
Aroclor-1016-4	5.856	5.745	5.945	522.540	500.000	4.5
Aroclor-1016-5	6.149	6.038	6.238	552.540	500.000	10.5
Aroclor-1260-1	7.268	7.156	7.356	611.390	500.000	22.3
Aroclor-1260-2	7.522	7.410	7.610	566.430	500.000	13.3
Aroclor-1260-3	7.881	7.768	7.968	512.070	500.000	2.4
Aroclor-1260-4	8.106	7.992	8.192	503.310	500.000	0.7
Aroclor-1260-5	8.425	8.311	8.511	528.360	500.000	5.7
Decachlorobiphenyl	10.244	10.125	10.325	53.400	50.000	6.8
Tetrachloro-m-xylene	4.521	4.412	4.612	51.030	50.000	2.1

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 04/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP071065.D Time Analyzed: 15:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.904	4.803	5.003	488.150	500.000	-2.4
Aroclor-1016-2	4.923	4.822	5.022	473.870	500.000	-5.2
Aroclor-1016-3	5.100	4.999	5.199	506.590	500.000	1.3
Aroclor-1016-4	5.142	5.041	5.241	495.880	500.000	-0.8
Aroclor-1016-5	5.356	5.255	5.455	469.850	500.000	-6.0
Aroclor-1260-1	6.392	6.292	6.492	515.960	500.000	3.2
Aroclor-1260-2	6.580	6.480	6.680	518.230	500.000	3.6
Aroclor-1260-3	6.734	6.634	6.834	524.780	500.000	5.0
Aroclor-1260-4	7.205	7.105	7.305	499.510	500.000	-0.1
Aroclor-1260-5	7.447	7.346	7.546	467.040	500.000	-6.6
Decachlorobiphenyl	8.859	8.759	8.959	45.690	50.000	-8.6
Tetrachloro-m-xylene	3.818	3.716	3.916	46.500	50.000	-7.0