

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL03 Date Analyzed: 04/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP071086.D Time Analyzed: 08:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.673	5.563	5.763	488.510	500.000	-2.3
Aroclor-1016-2	5.694	5.585	5.785	497.170	500.000	-0.6
Aroclor-1016-3	5.758	5.648	5.848	488.370	500.000	-2.3
Aroclor-1016-4	5.855	5.745	5.945	522.450	500.000	4.5
Aroclor-1016-5	6.146	6.038	6.238	561.130	500.000	12.2
Aroclor-1260-1	7.266	7.156	7.356	510.760	500.000	2.2
Aroclor-1260-2	7.519	7.410	7.610	539.740	500.000	7.9
Aroclor-1260-3	7.879	7.768	7.968	526.510	500.000	5.3
Aroclor-1260-4	8.103	7.992	8.192	492.820	500.000	-1.4
Aroclor-1260-5	8.424	8.311	8.511	533.780	500.000	6.8
Decachlorobiphenyl	10.243	10.125	10.325	51.790	50.000	3.6
Tetrachloro-m-xylene	4.519	4.412	4.612	50.620	50.000	1.2

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL03 Date Analyzed: 04/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP071086.D Time Analyzed: 08:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.903	4.803	5.003	532.240	500.000	6.4
Aroclor-1016-2	4.921	4.822	5.022	513.690	500.000	2.7
Aroclor-1016-3	5.098	4.999	5.199	543.720	500.000	8.7
Aroclor-1016-4	5.141	5.041	5.241	482.630	500.000	-3.5
Aroclor-1016-5	5.354	5.255	5.455	558.100	500.000	11.6
Aroclor-1260-1	6.390	6.292	6.492	567.470	500.000	13.5
Aroclor-1260-2	6.578	6.480	6.680	550.160	500.000	10.0
Aroclor-1260-3	6.732	6.634	6.834	517.750	500.000	3.6
Aroclor-1260-4	7.203	7.105	7.305	477.130	500.000	-4.6
Aroclor-1260-5	7.444	7.346	7.546	441.490	500.000	-11.7
Decachlorobiphenyl	8.855	8.759	8.959	49.020	50.000	-2.0
Tetrachloro-m-xylene	3.816	3.716	3.916	53.750	50.000	7.5