

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 04/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP071101.D Time Analyzed: 16:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.669	5.563	5.763	492.290	500.000	-1.5
Aroclor-1016-2	5.691	5.585	5.785	490.180	500.000	-2.0
Aroclor-1016-3	5.753	5.648	5.848	484.960	500.000	-3.0
Aroclor-1016-4	5.850	5.745	5.945	515.450	500.000	3.1
Aroclor-1016-5	6.143	6.038	6.238	556.720	500.000	11.3
Aroclor-1260-1	7.261	7.156	7.356	537.560	500.000	7.5
Aroclor-1260-2	7.514	7.410	7.610	547.200	500.000	9.4
Aroclor-1260-3	7.874	7.768	7.968	535.010	500.000	7.0
Aroclor-1260-4	8.100	7.992	8.192	506.940	500.000	1.4
Aroclor-1260-5	8.419	8.311	8.511	519.560	500.000	3.9
Decachlorobiphenyl	10.237	10.125	10.325	50.940	50.000	1.9
Tetrachloro-m-xylene	4.515	4.412	4.612	49.470	50.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 04/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP071101.D Time Analyzed: 16:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.803	5.003	473.240	500.000	-5.4
Aroclor-1016-2	4.921	4.822	5.022	462.730	500.000	-7.5
Aroclor-1016-3	5.098	4.999	5.199	490.560	500.000	-1.9
Aroclor-1016-4	5.140	5.041	5.241	438.060	500.000	-12.4
Aroclor-1016-5	5.354	5.255	5.455	504.300	500.000	0.9
Aroclor-1260-1	6.389	6.292	6.492	553.500	500.000	10.7
Aroclor-1260-2	6.578	6.480	6.680	534.230	500.000	6.8
Aroclor-1260-3	6.731	6.634	6.834	507.910	500.000	1.6
Aroclor-1260-4	7.203	7.105	7.305	470.680	500.000	-5.9
Aroclor-1260-5	7.444	7.346	7.546	436.910	500.000	-12.6
Decachlorobiphenyl	8.856	8.759	8.959	49.910	50.000	-0.2
Tetrachloro-m-xylene	3.816	3.716	3.916	45.230	50.000	-9.5