

Report of Analysis

| | | | | | |
|--------------------|------------------------------|----------|--------------------|-------|-----------|
| Client: | Tully Construction Co., Inc. | | Date Collected: | | |
| Project: | MTA Rockaway Park | | Date Received: | | |
| Client Sample ID: | PB167434BL | | SDG No.: | Q1698 | |
| Lab Sample ID: | PB167434BL | | Matrix: | SOIL | |
| Analytical Method: | SW8082A | | % Solid: | 100 | Decanted: |
| Sample Wt/Vol: | 30.02 | Units: g | Final Vol: | 10000 | uL |
| Soil Aliquot Vol: | | uL | Test: | PCB | |
| Extraction Type: | | | Injection Volume : | | |
| GPC Factor : | 1.0 | PH : | | | |
| Prep Method : | SW3541B | | | | |

| | | | | |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
| PP071092.D | 1 | 04/03/25 08:54 | 04/04/25 13:06 | PB167434 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|--------|-----------|----------|------------|-------------------|
| TARGETS | | | | | | |
| 12674-11-2 | Aroclor-1016 | 0.0039 | U | 0.0039 | 0.017 | mg/Kg |
| 11104-28-2 | Aroclor-1221 | 0.0040 | U | 0.0040 | 0.017 | mg/Kg |
| 11141-16-5 | Aroclor-1232 | 0.0037 | U | 0.0037 | 0.017 | mg/Kg |
| 53469-21-9 | Aroclor-1242 | 0.0040 | U | 0.0040 | 0.017 | mg/Kg |
| 12672-29-6 | Aroclor-1248 | 0.0059 | U | 0.0059 | 0.017 | mg/Kg |
| 11097-69-1 | Aroclor-1254 | 0.0032 | U | 0.0032 | 0.017 | mg/Kg |
| 37324-23-5 | Aroclor-1262 | 0.0050 | U | 0.0050 | 0.017 | mg/Kg |
| 11100-14-4 | Aroclor-1268 | 0.0036 | U | 0.0036 | 0.017 | mg/Kg |
| 11096-82-5 | Aroclor-1260 | 0.0032 | U | 0.0032 | 0.017 | mg/Kg |
| SURROGATES | | | | | | |
| 877-09-8 | Tetrachloro-m-xylene | 21.5 | | 32 - 144 | 108% | SPK: 20 |
| 2051-24-3 | Decachlorobiphenyl | 21.9 | | 32 - 175 | 110% | SPK: 20 |

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit