

Surrogate Summary

SDG No.: Q1698

Client: Tully Construction Co., Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP070916.D	PIBLK-PP070916.D	Tetrachloro-m-xylene	1	20	18.8	94		60	140
		Decachlorobiphenyl	1	20	19.4	97		60	140
		Tetrachloro-m-xylene	2	20	18.1	91		60	140
		Decachlorobiphenyl	2	20	23.4	117		60	140
I.BLK-PP071054.D	PIBLK-PP071054.D	Tetrachloro-m-xylene	1	20	19.9	100		60	140
		Decachlorobiphenyl	1	20	22.7	114		60	140
		Tetrachloro-m-xylene	2	20	23.0	115		60	140
		Decachlorobiphenyl	2	20	18.8	94		60	140
PB167434BS	PB167434BS	Tetrachloro-m-xylene	1	20	25.5	127		32	144
		Decachlorobiphenyl	1	20	27.4	137		32	175
		Tetrachloro-m-xylene	2	20	23.2	116		32	144
		Decachlorobiphenyl	2	20	25.1	126		32	175
Q1698-01	B-9	Tetrachloro-m-xylene	1	20	23.0	115		32	144
		Decachlorobiphenyl	1	20	22.7	114		32	175
		Tetrachloro-m-xylene	2	20	28.3	142		32	144
		Decachlorobiphenyl	2	20	20.9	104		32	175
Q1698-07	B-10	Tetrachloro-m-xylene	1	20	21.7	108		32	144
		Decachlorobiphenyl	1	20	21.6	108		32	175
		Tetrachloro-m-xylene	2	20	25.7	129		32	144
		Decachlorobiphenyl	2	20	18.3	91		32	175
Q1698-07MS	B-10MS	Tetrachloro-m-xylene	1	20	24.0	120		32	144
		Decachlorobiphenyl	1	20	23.3	116		32	175
		Tetrachloro-m-xylene	2	20	23.9	120		32	144
		Decachlorobiphenyl	2	20	20.6	103		32	175
Q1698-07MSD	B-10MSD	Tetrachloro-m-xylene	1	20	21.6	108		32	144
		Decachlorobiphenyl	1	20	23.1	116		32	175
		Tetrachloro-m-xylene	2	20	24.0	120		32	144
		Decachlorobiphenyl	2	20	20.8	104		32	175
I.BLK-PP071069.D	PIBLK-PP071069.D	Tetrachloro-m-xylene	1	20	20.4	102		60	140
		Decachlorobiphenyl	1	20	22.6	113		60	140
		Tetrachloro-m-xylene	2	20	22.0	110		60	140
		Decachlorobiphenyl	2	20	22.2	111		60	140
I.BLK-PP071090.D	PIBLK-PP071090.D	Tetrachloro-m-xylene	1	20	20.5	103		60	140
		Decachlorobiphenyl	1	20	21.6	108		60	140
		Tetrachloro-m-xylene	2	20	19.5	97		60	140
		Decachlorobiphenyl	2	20	20.5	103		60	140
PB167434BL	PB167434BL	Tetrachloro-m-xylene	1	20	21.5	108		32	144
		Decachlorobiphenyl	1	20	21.9	110		32	175
		Tetrachloro-m-xylene	2	20	20.3	102		32	144
		Decachlorobiphenyl	2	20	19.1	96		32	175
I.BLK-PP071105.D	PIBLK-PP071105.D	Tetrachloro-m-xylene	1	20	20.5	103		60	140

Surrogate Summary

SDG No.: Q1698

Client: Tully Construction Co., Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071105.D	PIBLK-PP071105.D	Decachlorobiphenyl	1	20	21.6	108		60	140
		Tetrachloro-m-xylene	2	20	20.1	100		60	140
		Decachlorobiphenyl	2	20	20.9	105		60	140