

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: TULL02
 ProjectID: MTA Rockaway Park
 Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG No.: Q1698

Calibration Sequence : FG032725		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	202510102	119124	FG015564.D
850	97431427	114625	FG015565.D
340	39672736	116685	FG015566.D
170	20516512	120685	FG015567.D
85	9074544	106759	FG015568.D
AVG RF : 115576		% RSD : 4.711	AVG RT : 15.021