

## Report of Analysis

Client:	Tully Construction Co., Inc.		Date Collected:	04/04/25	
Project:	MTA Rockaway Park		Date Received:	04/04/25	
Client Sample ID:	PIBLK-FG015635.D		SDG No.:	Q1698	
Lab Sample ID:	I.BLK-FG015635.D		Matrix:	Water	
Analytical Method:	8015D TPH		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	TPH GC	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3510				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG015635.D	1		04/04/25	FG040325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
PHC	Petroleum Hydrocarbons	0.012	U	0.012	0.085	mg/L
<b>SURROGATES</b>						
16416-32-3	TETRACOSANE-d50	16.9		29 - 130	84%	SPK: 20

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit