

SOIL TPH GC SURROGATE RECOVERY

Lab Name: Chemtech Client: Tully Construction Co., Inc.
Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG No.: Q1698

EPA SAMPLE NO.	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PIBLK-FG015612.D	82				0
PIBLK-FG015620.D	88				0
PIBLK-FG015629.D	84				0
PIBLK-FG015635.D	84				0
PB167438BL	94				0
PB167438BS	98				0
B-9	0 *				1
B-9-TPH-2	108				0
B-9-TPH-3	0 *				1
B-9-TPH-4	0 *				1
B-10	69				0
B-10MS	71				0
B-10MSD	71				0
B-10-TPH-2	69				0
B-10-TPH-3	79				0
B-10-TPH-4	76				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130
For Soil : 37-130

Column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogate Diluted Out