

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1698

Client: Tully Construction Co., Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1694-01MSD	Client Sample ID:	TP-18MSD					DataFile:	BM049823.D		
Naphthalene	1100	0	1000	ug/Kg	91		18		72	110	20
Acenaphthylene	1100	0	1200	ug/Kg	109		15		79	118	20
Acenaphthene	1100	0	1100	ug/Kg	100		17		70	121	20
Fluorene	1100	0	1100	ug/Kg	100		17		68	116	20
Phenanthrene	1100	0	1100	ug/Kg	100		17		52	128	20
Anthracene	1100	0	1200	ug/Kg	109		0		62	124	20
Fluoranthene	1100	0	1200	ug/Kg	109		9		44	125	20
Pyrene	1100	0	1100	ug/Kg	100		9		26	142	20
Benzo(a)anthracene	1100	0	1100	ug/Kg	100		24	*	71	114	20
Chrysene	1100	0	1100	ug/Kg	100		0		57	121	20
Benzo(b)fluoranthene	1100	0	1100	ug/Kg	100		17		67	121	20
Benzo(k)fluoranthene	1100	0	1100	ug/Kg	100		17		57	134	20
Benzo(a)pyrene	1100	0	1200	ug/Kg	109		0		70	142	20
Indeno(1,2,3-cd)pyrene	1100	0	1200	ug/Kg	109		9		40	129	20
Dibenz(a,h)anthracene	1100	0	1200	ug/Kg	109		9		43	123	20
Benzo(g,h,i)perylene	1100	0	1100	ug/Kg	100		0		24	125	20