

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
 Data File : VY021659.D
 Acq On : 27 Mar 2025 11:16
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 03/28/2025
 Supervised By :Mahesh Dadoda 03/28/2025

Quant Time: Mar 28 02:16:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:08:57 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	238234	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	369963	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	329626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	167924	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	130485	50.552	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 101.100%			
35) Dibromofluoromethane	7.634	113	125415	52.259	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 104.520%			
50) Toluene-d8	10.109	98	483058	52.909	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 105.820%			
62) 4-Bromofluorobenzene	12.407	95	162384	52.583	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 105.160%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	120589	48.836	ug/l	100
3) Chloromethane	2.068	50	164961	47.510	ug/l	100
4) Vinyl Chloride	2.202	62	190491	48.513	ug/l	100
5) Bromomethane	2.592	94	122288	45.662	ug/l	100
6) Chloroethane	2.732	64	122027	47.492	ug/l	100
7) Trichlorofluoromethane	3.055	101	241339	47.980	ug/l	100
8) Diethyl Ether	3.452	74	65605	47.960	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.818	101	127763	48.159	ug/l	100
10) Methyl Iodide	4.007	142	156113	53.639	ug/l	100
11) Tert butyl alcohol	4.860	59	40298	240.362	ug/l	100
12) 1,1-Dichloroethene	3.787	96	122422	48.991	ug/l	100
13) Acrolein	3.647	56	75366	246.189	ug/l	100
14) Allyl chloride	4.385	41	199515	50.077	ug/l	100
15) Acrylonitrile	5.061	53	144175	252.430	ug/l	100
16) Acetone	3.872	43	127702	241.737	ug/l	100
17) Carbon Disulfide	4.104	76	404128	49.101	ug/l	100
18) Methyl Acetate	4.385	43	63249	48.309	ug/l	100
19) Methyl tert-butyl Ether	5.116	73	315679	50.742	ug/l	100
20) Methylene Chloride	4.616	84	134344	52.136	ug/l	100
21) trans-1,2-Dichloroethene	5.116	96	135112	49.108	ug/l	100
22) Diisopropyl ether	6.018	45	428881	51.315	ug/l	100
23) Vinyl Acetate	5.957	43	1269853	259.718	ug/l	100
24) 1,1-Dichloroethane	5.915	63	247826	49.589	ug/l	100
25) 2-Butanone	6.896	43	189416	251.037	ug/l	100
26) 2,2-Dichloropropane	6.884	77	219131	49.293	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	155365	50.214	ug/l	100
28) Bromochloromethane	7.244	49	101413	48.116	ug/l	100
29) Tetrahydrofuran	7.262	42	125387	258.206	ug/l	100
30) Chloroform	7.421	83	257511	49.541	ug/l	100
31) Cyclohexane	7.701	56	218689	48.004	ug/l	100
32) 1,1,1-Trichloroethane	7.616	97	230441	49.051	ug/l	100
36) 1,1-Dichloropropene	7.835	75	185559	50.763	ug/l	100
37) Ethyl Acetate	6.988	43	87321	50.617	ug/l	100
38) Carbon Tetrachloride	7.823	117	208537	49.810	ug/l	100
39) Methylcyclohexane	9.109	83	229134	52.472	ug/l	100
40) Benzene	8.079	78	554462	50.649	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
 Data File : VY021659.D
 Acq On : 27 Mar 2025 11:16
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 03/28/2025
 Supervised By :Mahesh Dadoda 03/28/2025

Quant Time: Mar 28 02:16:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:08:57 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.232	41	50658m	55.298	ug/l	
42) 1,2-Dichloroethane	8.158	62	155365	50.343	ug/l	100
43) Isopropyl Acetate	8.195	43	166982	51.076	ug/l	100
44) Trichloroethene	8.865	130	138449	49.637	ug/l	100
45) 1,2-Dichloropropane	9.140	63	130586	50.395	ug/l	100
46) Dibromomethane	9.231	93	73435	49.174	ug/l	100
47) Bromodichloromethane	9.420	83	194399	50.237	ug/l	100
48) Methyl methacrylate	9.219	41	77877	51.902	ug/l	100
49) 1,4-Dioxane	9.225	88	16481	1030.192	ug/l	100
51) 4-Methyl-2-Pentanone	9.999	43	445564	261.893	ug/l	100
52) Toluene	10.170	92	351715	51.496	ug/l	100
53) t-1,3-Dichloropropene	10.396	75	179747	52.296	ug/l	100
54) cis-1,3-Dichloropropene	9.853	75	209143	51.579	ug/l	100
55) 1,1,2-Trichloroethane	10.572	97	96578	50.254	ug/l	100
56) Ethyl methacrylate	10.438	69	135272	55.124	ug/l	100
57) 1,3-Dichloropropane	10.719	76	170109	51.252	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.713	63	332369	277.643	ug/l	100
59) 2-Hexanone	10.761	43	302538	267.012	ug/l	100
60) Dibromochloromethane	10.908	129	134696	51.334	ug/l	100
61) 1,2-Dibromoethane	11.017	107	93866	52.067	ug/l	100
64) Tetrachloroethene	10.646	164	147971	49.288	ug/l	100
65) Chlorobenzene	11.444	112	380661	50.378	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.517	131	135660	50.602	ug/l	100
67) Ethyl Benzene	11.517	91	676205	52.188	ug/l	100
68) m/p-Xylenes	11.627	106	519154	104.634	ug/l	100
69) o-Xylene	11.956	106	243359	53.135	ug/l	100
70) Styrene	11.969	104	406215	53.195	ug/l	100
71) Bromoform	12.133	173	78206	51.140	ug/l	100
73) Isopropylbenzene	12.255	105	642393	51.505	ug/l	100
74) N-amyl acetate	12.072	43	154054	52.630	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.505	83	110292	49.185	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	83656m	50.257	ug/l	
77) Bromobenzene	12.529	156	149323	49.864	ug/l	100
78) n-propylbenzene	12.596	91	775252	51.702	ug/l	100
79) 2-Chlorotoluene	12.682	91	444178	51.001	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	528425	51.839	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	37240	51.550	ug/l	100
82) 4-Chlorotoluene	12.779	91	463024	50.995	ug/l	100
83) tert-Butylbenzene	12.999	119	467961	51.541	ug/l	100
84) 1,2,4-Trimethylbenzene	13.041	105	526387	52.248	ug/l	100
85) sec-Butylbenzene	13.176	105	687573	51.820	ug/l	100
86) p-Isopropyltoluene	13.291	119	576012	52.525	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	293534	49.752	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	292315	50.147	ug/l	100
89) n-Butylbenzene	13.615	91	532897	52.544	ug/l	100
90) Hexachloroethane	13.877	117	118708	49.396	ug/l	100
91) 1,2-Dichlorobenzene	13.657	146	255877	49.821	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	17123	49.268	ug/l	100
93) 1,2,4-Trichlorobenzene	14.919	180	147434	52.935	ug/l	100
94) Hexachlorobutadiene	15.023	225	88090	50.642	ug/l	100
95) Naphthalene	15.145	128	253209	54.848	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	125294	52.754	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
Data File : VY021659.D
Acq On : 27 Mar 2025 11:16
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 03/28/2025
Supervised By :Mahesh Dadoda 03/28/2025

Quant Time: Mar 28 02:16:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:08:57 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
 Data File : VY021659.D
 Acq On : 27 Mar 2025 11:16
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Quant Time: Mar 28 02:16:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:08:57 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Romaben Patel 03/28/2025
 Supervised By :Mahesh Dadoda 03/28/2025

