

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL02
Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698
Lab File ID: VY021764.D Date Analyzed: 04/03/2025
Instrument ID: MSVOA_Y Time Analyzed: 11:01
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	208234	7.71	319384	8.62	290244	11.42
UPPER LIMIT	416468	8.207	638768	9.116	580488	11.92
LOWER LIMIT	104117	7.207	159692	8.116	145122	10.92
EPA SAMPLE NO.						
B-9	254509	7.71	494707	8.62	458971	11.41
B-10	264094	7.71	502539	8.62	458507	11.41
VY0403SBL01	249432	7.71	475813	8.62	424056	11.41
VY0403SBS01	203899	7.71	319045	8.62	287712	11.42
VY0403SBSD01	206953	7.71	330176	8.62	290460	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.