

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL02
 Lab Code: CHEM Case No.: Q1698 SAS No.: Q1698 SDG NO.: Q1698
 Lab File ID: VY021764.D Date Analyzed: 04/03/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:01
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	149312	13.347				
UPPER LIMIT	298624	13.847				
LOWER LIMIT	74656	12.847				
EPA SAMPLE NO.						
B-9	190883	13.35				
B-10	179643	13.35				
VY0403SBL01	162342	13.35				
VY0403SBS01	148166	13.35				
VY0403SBSD01	151135	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.