

### Laboratory Control Sample Summary

**Client:** Aramark Uniforms

**SDG No.:** Q1703

**Project:** Monthly 2025

**Run No.:** LB135277

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB135277BS |            |        |                 |            |                 |                     |               |
| TSS       | mg/L       | 550        | 532    |                 | 97         | 1               | 90-110              | 04/03/2025    |

### Laboratory Control Sample Summary

**Client:** Aramark Uniforms

**SDG No.:** Q1703

**Project:** Monthly 2025

**Run No.:** LB135279

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB135279BS |            |        |                 |            |                 |                     |               |
| BOD5      | mg/L       | 198        | 188    |                 | 95         | 1               | 84.6-115.4          | 04/02/2025    |

### Laboratory Control Sample Summary

**Client:** Aramark Uniforms

**SDG No.:** Q1703

**Project:** Monthly 2025

**Run No.:** LB135297

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB135297BS |            |        |                 |            |                 |                     |               |
| TPH       | mg/L       | 20.0       | 17.0   |                 | 85         | 1               | 78-114              | 04/04/2025    |

### Laboratory Control Sample Summary

**Client:** Aramark Uniforms

**SDG No.:** Q1703

**Project:** Monthly 2025

**Run No.:** LB135297

| Analyte   | Units       | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|-------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB135297BSD |            |        |                 |            |                 |                     |               |
| TPH       | mg/L        | 20.0       | 17.2   |                 | 86         | 1               | 78-114              | 04/04/2025    |