



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms **SDG No.:** Q1703
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1703 **SAS No.:** Q1703
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.76	6.0	96	40 - 160	P	04/03/2025	12:20	LB135289
	Copper	22.8	20.0	114	40 - 160	P	04/03/2025	12:20	LB135289
	Lead	11.6	12.0	96	40 - 160	P	04/03/2025	12:20	LB135289
	Nickel	38.7	40.0	97	40 - 160	P	04/03/2025	12:20	LB135289
	Zinc	42.8	40.0	107	40 - 160	P	04/03/2025	12:20	LB135289
CRI01	Cadmium	5.43	6.0	90	40 - 160	P	04/09/2025	14:35	LB135361
	Copper	21.0	20.0	105	40 - 160	P	04/09/2025	14:35	LB135361
	Lead	11.5	12.0	96	40 - 160	P	04/09/2025	14:35	LB135361
	Nickel	37.0	40.0	92	40 - 160	P	04/09/2025	14:35	LB135361
	Zinc	37.6	40.0	94	40 - 160	P	04/09/2025	14:35	LB135361
CRA	Mercury	0.062	0.05	124	40 - 160	CV	04/15/2025	10:20	LB135432