

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms **SDG No.:** Q1703
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1703 **SAS No.:** Q1703

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167442BS							
Cadmium	ug/L	100	97.2		97	85 - 115	P
Copper	ug/L	150	158		105	85 - 115	P
Lead	ug/L	500	478		96	85 - 115	P
Nickel	ug/L	250	249		100	85 - 115	P
Zinc	ug/L	100	103		103	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q1703

Contract: ARAM01

Lab Code: CHEM

Case No.: Q1703

SAS No.: Q1703

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167601BS Mercury	ug/L	4.0	3.70		92	85 - 115	CV