



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q1704

Contract: PSEG04

Lab Code: CHEM

Case No.: Q1704

SAS No.: Q1704

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	12:15	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC		SDG No.: Q1704							
Contract: PSEG04	Lab Code: CHEM	Case No.: Q1704	SAS No.: Q1704						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	12:47	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
CCB02	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	13:57	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
CCB03	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	14:43	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
CCB04	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	15:36	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
CCB05	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	16:40	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
CCB06	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	18:00	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
CCB07	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	18:33	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q1704

Contract: PSEG04

Lab Code: CHEM

Case No.: Q1704

SAS No.: Q1704

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	14:30	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	14:30	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	14:30	LB135361

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC **SDG No.:** Q1704
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1704 **SAS No.:** Q1704

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	15:00	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	15:00	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	15:00	LB135361
CCB02	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	15:27	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	15:27	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	15:27	LB135361
CCB03	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	16:30	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	16:30	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	16:30	LB135361
CCB04	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	17:27	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	17:27	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	17:27	LB135361
CCB05	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	18:13	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	18:13	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	18:13	LB135361
CCB06	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	18:40	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	18:40	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	18:40	LB135361

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q1704

Contract: PSEG04

Lab Code: CHEM

Case No.: Q1704

SAS No.: Q1704

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------