

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Parkway Generation Operating LLC **SDG No.:** Q1704
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1704 **SAS No.:** Q1704
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	3.95	2.0	198	-18	22	04/03/2025	12:24	LB135289
	Nickel	1.10	2.0	55	-38	42	04/03/2025	12:24	LB135289
	Zinc	5.44			-40	40	04/03/2025	12:24	LB135289
ICSAB01	Copper	512	511	100	434	588	04/03/2025	12:28	LB135289
	Nickel	1010	954	106	810	1100	04/03/2025	12:28	LB135289
	Zinc	1070	952	112	809	1095	04/03/2025	12:28	LB135289
ICSA	Copper	10.4	2.0	522	-18	22	04/03/2025	12:33	LB135289
	Nickel	-1.42	2.0	71	-38	42	04/03/2025	12:33	LB135289
	Zinc	-5.86			-40	40	04/03/2025	12:33	LB135289
ICSAB	Copper	538	511	105	434	588	04/03/2025	12:37	LB135289
	Nickel	1010	954	106	810	1100	04/03/2025	12:37	LB135289
	Zinc	1070	952	112	809	1095	04/03/2025	12:37	LB135289
ICSA01	Copper	0.93	2.0	46	-18	22	04/09/2025	14:39	LB135361
	Nickel	2.38	2.0	119	-38	42	04/09/2025	14:39	LB135361
	Zinc	2.22			-40	40	04/09/2025	14:39	LB135361
ICSAB01	Copper	483	511	94	434	588	04/09/2025	14:43	LB135361
	Nickel	982	954	103	810	1100	04/09/2025	14:43	LB135361
	Zinc	1030	952	108	809	1095	04/09/2025	14:43	LB135361
ICSA	Copper	-2.16	2.0	108	-18	22	04/09/2025	14:48	LB135361
	Nickel	8.09	2.0	404	-38	42	04/09/2025	14:48	LB135361
	Zinc	9.91			-40	40	04/09/2025	14:48	LB135361
ICSAB	Copper	527	511	103	434	588	04/09/2025	14:52	LB135361
	Nickel	1010	954	106	810	1100	04/09/2025	14:52	LB135361
	Zinc	1030	952	108	809	1095	04/09/2025	14:52	LB135361