

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Parkway Generation Operating LLC **level:** low **sdg no.:** Q1704
contract: PSEG04 **lab code:** CHEM **case no.:** Q1704 **sas no.:** Q1704
matrix: Water **sample id:** Q1689-01 **client id:** Q2 WASTEWATERMS
Percent Solids for Sample: NA **Spiked ID:** Q1689-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	129	D	100	UD	150	82		P
Nickel	ug/L	75 - 125	1630	D	1450	D	250	71		P
Zinc	ug/L	75 - 125	204	D	86.2	JD	100	118		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Parkway Generation Operating LLC **level:** low **sdg no.:** Q1704
contract: PSEG04 **lab code:** CHEM **case no.:** Q1704 **sas no.:** Q1704
matrix: Water **sample id:** Q1689-01 **client id:** Q2 WASTEWATERMSD
Percent Solids for Sample: NA **Spiked ID:** Q1689-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	131	D	100	UD	150	83		P
Nickel	ug/L	75 - 125	1640	D	1450	D	250	77		P
Zinc	ug/L	75 - 125	201	D	86.2	JD	100	115		P