

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1707

Project: Transfer Station - SPDES 2025

Run No.: LB135277

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135277BS							
TSS	mg/L	550	532		97	1	90-110	04/03/2025