

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1707
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1707 **SAS No.:** Q1707
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	516	510	101	95 - 105	P	04/03/2025	12:06	LB135289
	Iron	10300	10000	103	95 - 105	P	04/03/2025	12:06	LB135289
	Lead	1010	1000	101	95 - 105	P	04/03/2025	12:06	LB135289

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1707
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1707 **SAS No.:** Q1707
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.8	20.0	114	80 - 120	P	04/03/2025	12:11	LB135289
	Iron	102	100	102	80 - 120	P	04/03/2025	12:11	LB135289
	Lead	11.1	12.0	93	80 - 120	P	04/03/2025	12:11	LB135289

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1707
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1707 **SAS No.:** Q1707
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1270	1250	102	90 - 110	P	04/03/2025	12:41	LB135289
	Iron	4870	5000	97	90 - 110	P	04/03/2025	12:41	LB135289
	Lead	4930	5000	99	90 - 110	P	04/03/2025	12:41	LB135289
CCV02	Copper	1260	1250	101	90 - 110	P	04/03/2025	13:53	LB135289
	Iron	4860	5000	97	90 - 110	P	04/03/2025	13:53	LB135289
	Lead	4860	5000	97	90 - 110	P	04/03/2025	13:53	LB135289
CCV03	Copper	1270	1250	102	90 - 110	P	04/03/2025	14:39	LB135289
	Iron	4880	5000	98	90 - 110	P	04/03/2025	14:39	LB135289
	Lead	4910	5000	98	90 - 110	P	04/03/2025	14:39	LB135289
CCV04	Copper	1250	1250	100	90 - 110	P	04/03/2025	15:32	LB135289
	Iron	4970	5000	99	90 - 110	P	04/03/2025	15:32	LB135289
	Lead	4850	5000	97	90 - 110	P	04/03/2025	15:32	LB135289
CCV05	Copper	1250	1250	100	90 - 110	P	04/03/2025	16:36	LB135289
	Iron	4900	5000	98	90 - 110	P	04/03/2025	16:36	LB135289
	Lead	4860	5000	97	90 - 110	P	04/03/2025	16:36	LB135289
CCV06	Copper	1280	1250	103	90 - 110	P	04/03/2025	17:48	LB135289
	Iron	4880	5000	98	90 - 110	P	04/03/2025	17:48	LB135289
	Lead	4940	5000	99	90 - 110	P	04/03/2025	17:48	LB135289
CCV07	Copper	1280	1250	103	90 - 110	P	04/03/2025	18:29	LB135289
	Iron	4960	5000	99	90 - 110	P	04/03/2025	18:29	LB135289
	Lead	4980	5000	100	90 - 110	P	04/03/2025	18:29	LB135289

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1707
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1707 **SAS No.:** Q1707
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	515	510	101	95 - 105	P	04/09/2025	14:21	LB135361
	Iron	9500	10000	95	95 - 105	P	04/09/2025	14:21	LB135361
	Lead	1020	1000	102	95 - 105	P	04/09/2025	14:21	LB135361

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1707
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1707 **SAS No.:** Q1707
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.5	20.0	108	80 - 120	P	04/09/2025	14:26	LB135361
	Iron	96.6	100	97	80 - 120	P	04/09/2025	14:26	LB135361
	Lead	11.7	12.0	97	80 - 120	P	04/09/2025	14:26	LB135361

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1707
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1707 **SAS No.:** Q1707
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1230	1250	98	90 - 110	P	04/09/2025	14:56	LB135361
	Iron	5030	5000	100	90 - 110	P	04/09/2025	14:56	LB135361
	Lead	4870	5000	97	90 - 110	P	04/09/2025	14:56	LB135361
CCV02	Copper	1210	1250	97	90 - 110	P	04/09/2025	15:23	LB135361
	Iron	4840	5000	97	90 - 110	P	04/09/2025	15:23	LB135361
	Lead	4730	5000	95	90 - 110	P	04/09/2025	15:23	LB135361
CCV03	Copper	1220	1250	98	90 - 110	P	04/09/2025	16:26	LB135361
	Iron	4920	5000	98	90 - 110	P	04/09/2025	16:26	LB135361
	Lead	4790	5000	96	90 - 110	P	04/09/2025	16:26	LB135361
CCV04	Copper	1240	1250	99	90 - 110	P	04/09/2025	17:23	LB135361
	Iron	5030	5000	101	90 - 110	P	04/09/2025	17:23	LB135361
	Lead	4880	5000	98	90 - 110	P	04/09/2025	17:23	LB135361
CCV05	Copper	1240	1250	100	90 - 110	P	04/09/2025	18:09	LB135361
	Iron	4980	5000	100	90 - 110	P	04/09/2025	18:09	LB135361
	Lead	4890	5000	98	90 - 110	P	04/09/2025	18:09	LB135361
CCV06	Copper	1200	1250	96	90 - 110	P	04/09/2025	18:36	LB135361
	Iron	4930	5000	99	90 - 110	P	04/09/2025	18:36	LB135361
	Lead	4680	5000	94	90 - 110	P	04/09/2025	18:36	LB135361