

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1707
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1707 **SAS No.:** Q1707
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	3.95	2.0	198	-18	22	04/03/2025	12:24	LB135289
	Iron	93800	101000	93	85600	116500	04/03/2025	12:24	LB135289
	Lead	4.36			-12	12	04/03/2025	12:24	LB135289
ICSAB01	Copper	512	511	100	434	588	04/03/2025	12:28	LB135289
	Iron	99000	99300	100	84400	114500	04/03/2025	12:28	LB135289
	Lead	55.2	49.0	113	37	61	04/03/2025	12:28	LB135289
ICSA	Copper	10.4	2.0	522	-18	22	04/03/2025	12:33	LB135289
	Iron	90700	101000	90	85600	116500	04/03/2025	12:33	LB135289
	Lead	8.69			-12	12	04/03/2025	12:33	LB135289
ICSAB	Copper	538	511	105	434	588	04/03/2025	12:37	LB135289
	Iron	102000	99300	103	84400	114500	04/03/2025	12:37	LB135289
	Lead	56.0	49.0	114	37	61	04/03/2025	12:37	LB135289
ICSA01	Copper	0.93	2.0	46	-18	22	04/09/2025	14:39	LB135361
	Iron	97200	101000	96	85600	116500	04/09/2025	14:39	LB135361
	Lead	6.82			-12	12	04/09/2025	14:39	LB135361
ICSAB01	Copper	483	511	94	434	588	04/09/2025	14:43	LB135361
	Iron	106000	99300	107	84400	114500	04/09/2025	14:43	LB135361
	Lead	51.3	49.0	105	37	61	04/09/2025	14:43	LB135361
ICSA	Copper	-2.16	2.0	108	-18	22	04/09/2025	14:48	LB135361
	Iron	105000	101000	104	85600	116500	04/09/2025	14:48	LB135361
	Lead	16.4			-12	12	04/09/2025	14:48	LB135361
ICSAB	Copper	527	511	103	434	588	04/09/2025	14:52	LB135361
	Iron	97000	99300	98	84400	114500	04/09/2025	14:52	LB135361
	Lead	52.0	49.0	106	37	61	04/09/2025	14:52	LB135361