

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q1707

Contract: TULL01

Lab Code: CHEM

Case No.: Q1707

SAS No.: Q1707

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167442BS							
Copper	ug/L	150	158		105	85 - 115	P
Iron	ug/L	1500	1480		99	85 - 115	P
Lead	ug/L	500	478		96	85 - 115	P