

**ALLIANCE** 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: Q1711

MATRIX: Water

METHOD: 9056A,SM2320 B,SM2540 C

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all samples. The Matrix Spike (MW-18B-56-040225MS) analysis met criteria for all samples except for Chloride due to matrix interference. The Matrix Spike Duplicate (MW-18B-56-040225MSD) analysis met criteria for all samples except for due to matrix interference.			
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

\_\_\_\_\_  
QA REVIEW

\_\_\_\_\_  
Date