

## Duplicate Sample Summary

|                                                           |                                           |
|-----------------------------------------------------------|-------------------------------------------|
| <b>Client:</b> JACOBS Engineering Group, Inc.             | <b>SDG No.:</b> Q1711                     |
| <b>Project:</b> Former Schlumberger STC PTC Site D3868221 | <b>Sample ID:</b> Q1711-01                |
| <b>Client ID:</b> MW-18B-56-040225DUP                     | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte    | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|------------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| TDS        | mg/L  | +/-5             | 257           |                 | 261              |                 | 1               | 1.54   |      | 04/03/2025    |
| Alkalinity | mg/L  | +/-20            | 273           |                 | 270              |                 | 1               | 1      |      | 04/03/2025    |

## Duplicate Sample Summary

|                                                                                                                                                     |                                                                                                  |
|-----------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------|
| <b>Client:</b> JACOBS Engineering Group, Inc.<br><b>Project:</b> Former Schlumberger STC PTC Site D3868221<br><b>Client ID:</b> MW-18B-56-040225MSD | <b>SDG No.:</b> Q1711<br><b>Sample ID:</b> Q1711-01<br><b>Percent Solids for Spike Sample:</b> 0 |
|-----------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------|

| Analyte             | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|---------------------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| Chloride            | mg/L  | +/-15            | 28.9          | OR              | 28.9             | OR              | 1               | 0      |      | 04/03/2025    |
| Sulfate             | mg/L  | +/-15            | 24.9          |                 | 24.5             |                 | 1               | 2      |      | 04/03/2025    |
| Bromide             | mg/L  | +/-15            | 10.4          |                 | 10.1             |                 | 1               | 3      |      | 04/03/2025    |
| Nitrite             | mg/L  | +/-15            | 3.10          |                 | 3.00             |                 | 1               | 3      |      | 04/03/2025    |
| Fluoride            | mg/L  | +/-15            | 2.30          |                 | 2.20             |                 | 1               | 4      |      | 04/03/2025    |
| Nitrate             | mg/L  | +/-15            | 2.60          |                 | 2.50             |                 | 1               | 4      |      | 04/03/2025    |
| Orthophosphate as P | mg/L  | +/-15            | 2.40          |                 | 1.60             |                 | 1               | 40     | *    | 04/03/2025    |

### Duplicate Sample Summary

|                   |                                           |                                         |          |
|-------------------|-------------------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | JACOBS Engineering Group, Inc.            | <b>SDG No.:</b>                         | Q1711    |
| <b>Project:</b>   | Former Schlumberger STC PTC Site D3868221 | <b>Sample ID:</b>                       | Q1711-08 |
| <b>Client ID:</b> | EB01-040225DUP                            | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte    | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Alkalinity | mg/L  | +/-20               | 1.00             | U                  | 1.00                | U                  | 1                  | 0          |      | 04/03/2025       |