

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1711

Project: Former Schlumberger STC PTC Site D3868221

Run No.: LB135286

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135286BS							
TDS	mg/L	100	94.0		94	1	90-110	04/03/2025

Laboratory Control Sample Summary

Client:	JACOBS Engineering Group, Inc.	SDG No.:	Q1711
Project:	Former Schlumberger STC PTC Site D3868221	Run No.:	LB135288

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135288BSW							
Alkalinity	mg/L	50	53.2		106	1	80-120	04/03/2025

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1711

Project: Former Schlumberger STC PTC Site D3868221

Run No.: LB135291

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135291BS							
Alkalinity	mg/L	50	45.4		91	1	80-120	04/03/2025

Laboratory Control Sample Summary

Client:	JACOBS Engineering Group, Inc.	SDG No.:	Q1711
Project:	Former Schlumberger STC PTC Site D3868221	Run No.:	LB135296

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135296BSW							
Bromide	mg/L	10	10.4		104	1	90-110	04/03/2025
Chloride	mg/L	3	3.10		103	1	90-110	04/03/2025
Fluoride	mg/L	2	2.00		100	1	90-110	04/03/2025
Nitrite	mg/L	3	3.10		103	1	90-110	04/03/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	04/03/2025
Sulfate	mg/L	15	15.2		101	1	90-110	04/03/2025
Orthophosphate as P	mg/L	5	5.20		104	1	90-110	04/03/2025