

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1711

Contract: JACO05

Lab Code: CHEM

Case No.: Q1711

SAS No.: Q1711

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167465BS							
Aluminum	ug/L	10000	10100		101	80 - 120	P
Antimony	ug/L	500	507		101	80 - 120	P
Arsenic	ug/L	500	511		102	80 - 120	P
Barium	ug/L	2500	2570		103	80 - 120	P
Beryllium	ug/L	500	511		102	80 - 120	P
Cadmium	ug/L	500	509		102	80 - 120	P
Chromium	ug/L	500	511		102	80 - 120	P
Copper	ug/L	5000	5130		103	80 - 120	P
Iron	ug/L	25000	27400		110	80 - 120	P
Lead	ug/L	2500	2520		101	80 - 120	P
Magnesium	ug/L	50000	51000		102	80 - 120	P
Manganese	ug/L	5000	5010		100	80 - 120	P
Potassium	ug/L	25000	25400		102	80 - 120	P
Selenium	ug/L	500	508		102	80 - 120	P
Sodium	ug/L	50000	52300		105	80 - 120	P