

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JACO05

Lab Code: CHEM Case no.: Q1711

Sas No.: Q1711 SDG No.: Q1711

Instrument ID: P7

Start Date : 04/04/2025

Run Number: LB135315

End Date : 04/04/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                  |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-----------------------------------------------|---|------------------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>165Ho                              | Q | Element<br>209Bi | Q | Element | Q | Element | Q | Element | Q |
| S0              | S0                 | 1409 | 100                                           |   | 100              |   |         |   |         |   |         |   |
| S2              | S2                 | 1415 | 99                                            |   | 99               |   |         |   |         |   |         |   |
| S3              | S3                 | 1419 | 99                                            |   | 101              |   |         |   |         |   |         |   |
| S4              | S4                 | 1422 | 100                                           |   | 101              |   |         |   |         |   |         |   |
| S5              | S5                 | 1425 | 99                                            |   | 103              |   |         |   |         |   |         |   |
| S6              | S6                 | 1428 | 101                                           |   | 103              |   |         |   |         |   |         |   |
| S7              | S7                 | 1431 | 104                                           |   | 102              |   |         |   |         |   |         |   |
| S8              | S8                 | 1433 | 114                                           |   | 105              |   |         |   |         |   |         |   |
| ICV01           | ICV01              | 1458 | 106                                           |   | 107              |   |         |   |         |   |         |   |
| LLICV           | LLICV              | 1517 | 105                                           |   | 107              |   |         |   |         |   |         |   |
| ICB01           | ICB01              | 1520 | 107                                           |   | 107              |   |         |   |         |   |         |   |
| ICSA01          | ICSA01             | 1524 | 109                                           |   | 105              |   |         |   |         |   |         |   |
| ICSAB01         | ICSAB01            | 1527 | 111                                           |   | 109              |   |         |   |         |   |         |   |
| CCV01           | CCV01              | 1530 | 112                                           |   | 106              |   |         |   |         |   |         |   |
| CCB01           | CCB01              | 1533 | 109                                           |   | 110              |   |         |   |         |   |         |   |
| CRI             | CRI                | 1537 | 109                                           |   | 111              |   |         |   |         |   |         |   |
| PB167465BL      | PB167465BL         | 1540 | 108                                           |   | 110              |   |         |   |         |   |         |   |
| PB167465BS      | PB167465BS         | 1543 | 111                                           |   | 110              |   |         |   |         |   |         |   |
| Q1711-01        | MW-18B-56-04       | 1546 | 122                                           |   | 121              |   |         |   |         |   |         |   |
| Q1711-01DUP     | MW-18B-56-04       | 1549 | 124                                           |   | 121              |   |         |   |         |   |         |   |
| Q1711-01L       | MW-18B-56-04       | 1553 | 112                                           |   | 112              |   |         |   |         |   |         |   |
| Q1711-02        | MW-18B-56-04       | 1556 | 124                                           |   | 122              |   |         |   |         |   |         |   |
| Q1711-03        | MW-18B-56-04       | 1559 | 123                                           |   | 122              |   |         |   |         |   |         |   |
| Q1711-01A       | MW-18B-56-04       | 1602 | 125                                           |   | 123              |   |         |   |         |   |         |   |
| CCV02           | CCV02              | 1610 | 116                                           |   | 111              |   |         |   |         |   |         |   |
| CCB02           | CCB02              | 1615 | 110                                           |   | 111              |   |         |   |         |   |         |   |
| Q1711-04        | MW-17B-55-04       | 1618 | 125                                           |   | 125              |   |         |   |         |   |         |   |
| Q1711-08        | EB01-040225        | 1621 | 124                                           |   | 126              |   |         |   |         |   |         |   |
| CCV03           | CCV03              | 1624 | 111                                           |   | 106              |   |         |   |         |   |         |   |
| CCB03           | CCB03              | 1628 | 110                                           |   | 112              |   |         |   |         |   |         |   |

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JACO05

Lab Code: CHEM Case no.: Q1711

Sas No.: Q1711 SDG No.: Q1711

Instrument ID: P7

Start Date : 04/04/2025

Run Number: LB135315

End Date : 04/04/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |         |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-------------------------------------------|---|---------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>209Bi                          | Q | Element | Q | Element | Q | Element | Q | Element | Q |
| S0              | S0                 | 1409 | 100                                       |   |         |   |         |   |         |   |         |   |
| S2              | S2                 | 1415 | 100                                       |   |         |   |         |   |         |   |         |   |
| S3              | S3                 | 1419 | 103                                       |   |         |   |         |   |         |   |         |   |
| S4              | S4                 | 1422 | 103                                       |   |         |   |         |   |         |   |         |   |
| S5              | S5                 | 1425 | 105                                       |   |         |   |         |   |         |   |         |   |
| S6              | S6                 | 1428 | 104                                       |   |         |   |         |   |         |   |         |   |
| S7              | S7                 | 1431 | 104                                       |   |         |   |         |   |         |   |         |   |
| S8              | S8                 | 1433 | 98                                        |   |         |   |         |   |         |   |         |   |
| ICV01           | ICV01              | 1458 | 107                                       |   |         |   |         |   |         |   |         |   |
| LLICV           | LLICV              | 1517 | 104                                       |   |         |   |         |   |         |   |         |   |
| ICB01           | ICB01              | 1520 | 108                                       |   |         |   |         |   |         |   |         |   |
| ICSA01          | ICSA01             | 1524 | 105                                       |   |         |   |         |   |         |   |         |   |
| ICSAB01         | ICSAB01            | 1527 | 109                                       |   |         |   |         |   |         |   |         |   |
| CCV01           | CCV01              | 1530 | 102                                       |   |         |   |         |   |         |   |         |   |
| CCB01           | CCB01              | 1533 | 109                                       |   |         |   |         |   |         |   |         |   |
| CRI             | CRI                | 1537 | 109                                       |   |         |   |         |   |         |   |         |   |
| PB167465BL      | PB167465BL         | 1540 | 109                                       |   |         |   |         |   |         |   |         |   |
| PB167465BS      | PB167465BS         | 1543 | 107                                       |   |         |   |         |   |         |   |         |   |
| Q1711-01        | MW-18B-56-04       | 1546 | 118                                       |   |         |   |         |   |         |   |         |   |
| Q1711-01DUP     | MW-18B-56-04       | 1549 | 117                                       |   |         |   |         |   |         |   |         |   |
| Q1711-01L       | MW-18B-56-04       | 1553 | 110                                       |   |         |   |         |   |         |   |         |   |
| Q1711-02        | MW-18B-56-04       | 1556 | 116                                       |   |         |   |         |   |         |   |         |   |
| Q1711-03        | MW-18B-56-04       | 1559 | 116                                       |   |         |   |         |   |         |   |         |   |
| Q1711-01A       | MW-18B-56-04       | 1602 | 117                                       |   |         |   |         |   |         |   |         |   |
| CCV02           | CCV02              | 1610 | 107                                       |   |         |   |         |   |         |   |         |   |
| CCB02           | CCB02              | 1615 | 109                                       |   |         |   |         |   |         |   |         |   |
| Q1711-04        | MW-17B-55-04       | 1618 | 122                                       |   |         |   |         |   |         |   |         |   |
| Q1711-08        | EB01-040225        | 1621 | 121                                       |   |         |   |         |   |         |   |         |   |
| CCV03           | CCV03              | 1624 | 104                                       |   |         |   |         |   |         |   |         |   |
| CCB03           | CCB03              | 1628 | 110                                       |   |         |   |         |   |         |   |         |   |

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JACO05

Lab Code: CHEM Case no.: Q1711

Sas No.: Q1711 SDG No.: Q1711

Instrument ID: P8

Start Date : 04/07/2025

Run Number: LB135332

End Date : 04/07/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                  |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-----------------------------------------------|---|------------------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>165Ho                              | Q | Element<br>209Bi | Q | Element | Q | Element | Q | Element | Q |
| S0              | S0                 | 1129 | 100                                           |   | 100              |   |         |   |         |   |         |   |
| S2              | S2                 | 1136 | 99                                            |   | 99               |   |         |   |         |   |         |   |
| S3              | S3                 | 1139 | 98                                            |   | 97               |   |         |   |         |   |         |   |
| S4              | S4                 | 1142 | 99                                            |   | 96               |   |         |   |         |   |         |   |
| S5              | S5                 | 1145 | 99                                            |   | 95               |   |         |   |         |   |         |   |
| S6              | S6                 | 1148 | 95                                            |   | 90               |   |         |   |         |   |         |   |
| S7              | S7                 | 1151 | 96                                            |   | 88               |   |         |   |         |   |         |   |
| S8              | S8                 | 1154 | 86                                            |   | 74               |   |         |   |         |   |         |   |
| ICV01           | ICV01              | 1240 | 101                                           |   | 100              |   |         |   |         |   |         |   |
| LLICV           | LLICV              | 1244 | 99                                            |   | 99               |   |         |   |         |   |         |   |
| ICB01           | ICB01              | 1248 | 98                                            |   | 99               |   |         |   |         |   |         |   |
| ICSA01          | ICSA01             | 1257 | 95                                            |   | 86               |   |         |   |         |   |         |   |
| ICSAB01         | ICSAB01            | 1304 | 96                                            |   | 88               |   |         |   |         |   |         |   |
| CCV01           | CCV01              | 1307 | 91                                            |   | 80               |   |         |   |         |   |         |   |
| CCB01           | CCB01              | 1317 | 101                                           |   | 101              |   |         |   |         |   |         |   |
| CRI             | CRI                | 1320 | 102                                           |   | 102              |   |         |   |         |   |         |   |
| Q1711-12        | MW-17B-55-04       | 1324 | 102                                           |   | 99               |   |         |   |         |   |         |   |
| Q1711-13        | EB01-040225        | 1327 | 102                                           |   | 101              |   |         |   |         |   |         |   |
| Q1711-14        | MW-18B-56-04       | 1330 | 99                                            |   | 95               |   |         |   |         |   |         |   |
| Q1711-14DUP     | MW-18B-56-04       | 1333 | 102                                           |   | 96               |   |         |   |         |   |         |   |
| Q1711-14L       | MW-18B-56-04       | 1337 | 100                                           |   | 97               |   |         |   |         |   |         |   |
| Q1711-15        | MW-18B-56-04       | 1340 | 99                                            |   | 93               |   |         |   |         |   |         |   |
| Q1711-16        | MW-18B-56-04       | 1343 | 100                                           |   | 91               |   |         |   |         |   |         |   |
| CCV02           | CCV02              | 1358 | 88                                            |   | 80               |   |         |   |         |   |         |   |
| CCB02           | CCB02              | 1405 | 97                                            |   | 97               |   |         |   |         |   |         |   |
| Q1711-14A       | MW-18B-56-04       | 1408 | 95                                            |   | 89               |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1411 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1414 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1418 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1421 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1433 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1436 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1439 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1443 |                                               |   |                  |   |         |   |         |   |         |   |

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JACO05

Lab Code: CHEM Case no.: Q1711

Sas No.: Q1711 SDG No.: Q1711

Instrument ID: P8

Start Date : 04/07/2025

Run Number: LB135332

End Date : 04/07/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                  |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-----------------------------------------------|---|------------------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>165Ho                              | Q | Element<br>209Bi | Q | Element | Q | Element | Q | Element | Q |
| ZZZZZZ          | ZZZZZZ             | 1452 |                                               |   |                  |   |         |   |         |   |         |   |
| CCV03           | CCV03              | 1500 | 93                                            |   | 83               |   |         |   |         |   |         |   |
| CCB03           | CCB03              | 1513 | 101                                           |   | 102              |   |         |   |         |   |         |   |

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JAC005

Lab Code: CHEM Case no.: Q1711

Sas No.: Q1711 SDG No.: Q1711

Instrument ID: P8

Start Date : 04/07/2025

Run Number: LB135332

End Date : 04/07/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |         |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-------------------------------------------|---|---------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>209Bi                          | Q | Element | Q | Element | Q | Element | Q | Element | Q |
| S0              | S0                 | 1129 | 100                                       |   |         |   |         |   |         |   |         |   |
| S2              | S2                 | 1136 | 101                                       |   |         |   |         |   |         |   |         |   |
| S3              | S3                 | 1139 | 99                                        |   |         |   |         |   |         |   |         |   |
| S4              | S4                 | 1142 | 97                                        |   |         |   |         |   |         |   |         |   |
| S5              | S5                 | 1145 | 95                                        |   |         |   |         |   |         |   |         |   |
| S6              | S6                 | 1148 | 92                                        |   |         |   |         |   |         |   |         |   |
| S7              | S7                 | 1151 | 89                                        |   |         |   |         |   |         |   |         |   |
| S8              | S8                 | 1154 | 76                                        |   |         |   |         |   |         |   |         |   |
| ICV01           | ICV01              | 1240 | 98                                        |   |         |   |         |   |         |   |         |   |
| LLICV           | LLICV              | 1244 | 98                                        |   |         |   |         |   |         |   |         |   |
| ICB01           | ICB01              | 1248 | 97                                        |   |         |   |         |   |         |   |         |   |
| ICSA01          | ICSA01             | 1257 | 86                                        |   |         |   |         |   |         |   |         |   |
| ICSAB01         | ICSAB01            | 1304 | 86                                        |   |         |   |         |   |         |   |         |   |
| CCV01           | CCV01              | 1307 | 81                                        |   |         |   |         |   |         |   |         |   |
| CCB01           | CCB01              | 1317 | 101                                       |   |         |   |         |   |         |   |         |   |
| CRI             | CRI                | 1320 | 101                                       |   |         |   |         |   |         |   |         |   |
| Q1711-12        | MW-17B-55-04       | 1324 | 99                                        |   |         |   |         |   |         |   |         |   |
| Q1711-13        | EB01-040225        | 1327 | 101                                       |   |         |   |         |   |         |   |         |   |
| Q1711-14        | MW-18B-56-04       | 1330 | 96                                        |   |         |   |         |   |         |   |         |   |
| Q1711-14DUP     | MW-18B-56-04       | 1333 | 99                                        |   |         |   |         |   |         |   |         |   |
| Q1711-14L       | MW-18B-56-04       | 1337 | 98                                        |   |         |   |         |   |         |   |         |   |
| Q1711-15        | MW-18B-56-04       | 1340 | 92                                        |   |         |   |         |   |         |   |         |   |
| Q1711-16        | MW-18B-56-04       | 1343 | 93                                        |   |         |   |         |   |         |   |         |   |
| CCV02           | CCV02              | 1358 | 83                                        |   |         |   |         |   |         |   |         |   |
| CCB02           | CCB02              | 1405 | 96                                        |   |         |   |         |   |         |   |         |   |
| Q1711-14A       | MW-18B-56-04       | 1408 | 89                                        |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1411 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1414 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1418 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1421 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1433 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1436 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1439 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1443 |                                           |   |         |   |         |   |         |   |         |   |

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JACO05

Lab Code: CHEM Case no.: Q1711

Sas No.: Q1711 SDG No.: Q1711

Instrument ID: P8

Start Date : 04/07/2025

Run Number: LB135332

End Date : 04/07/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |         |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-------------------------------------------|---|---------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>209Bi                          | Q | Element | Q | Element | Q | Element | Q | Element | Q |
| ZZZZZZ          | ZZZZZZ             | 1452 |                                           |   |         |   |         |   |         |   |         |   |
| CCV03           | CCV03              | 1500 | 84                                        |   |         |   |         |   |         |   |         |   |
| CCB03           | CCB03              | 1513 | 104                                       |   |         |   |         |   |         |   |         |   |