

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1711

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1711-03MSD	Client Sample ID:	MW-18B-56-040225MSD					DataFile:	BN036835.D		
1,4-Dioxane	0.41	4.00	3.70	ug/L	-73	*	29	*	20 (10)	160 (175)	20 (20)