

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1711 SAS No.: Q1711 SDG NO.: Q1711
 Lab File ID: VX045559.D Date Analyzed: 04/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:30
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	95221	5.54	162684	6.76	143783	10.05
UPPER LIMIT	190442	6.044	325368	7.257	287566	10.549
LOWER LIMIT	47610.5	5.044	81342	6.257	71891.5	9.549
EPA SAMPLE NO.						
MW-18B-56-040225	64116	5.55	121192	6.76	112037	10.06
MW-18B-56-040225MSD	69182	5.55	124452	6.76	109108	10.06
MW-17B-55-040225	63215	5.55	122949	6.76	111016	10.06
RMW-05B-89-040225	63845	5.55	125144	6.76	116415	10.05
EB01-040225	63287	5.54	123377	6.76	114226	10.05
TB01-040225	66937	5.54	129076	6.76	121728	10.05
VX0403WBL01	66309	5.55	126206	6.76	118586	10.05
VX0403WBS01	95034	5.54	168178	6.76	149052	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.