

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05  
 Lab Code: CHEM Case No.: Q1711 SAS No.: Q1711 SDG NO.: Q1711  
 Lab File ID: VX045586.D Date Analyzed: 04/04/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 10:00  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                    | IS4<br>AREA # | RT #   |  |  |  |  |
|--------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD        | 67379         | 12.018 |  |  |  |  |
| UPPER LIMIT        | 134758        | 12.518 |  |  |  |  |
| LOWER LIMIT        | 33689.5       | 11.518 |  |  |  |  |
| EPA SAMPLE NO.     |               |        |  |  |  |  |
| MW-18B-56-040225MS | 64924         | 12.02  |  |  |  |  |
| VX0404WBL01        | 46025         | 12.02  |  |  |  |  |
| VX0404WBS01        | 66584         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.