

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135301

Review By	Niha	Review On	4/4/2025 12:55:28 PM
Supervise By	Iwona	Supervise On	4/4/2025 1:30:18 PM
SubDirectory	LB135301	Test	Residual Chlorine
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP112595,WP112589,WP112590,WP112591,WP112588,WP112592,WP112593,W3147		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	04/04/25 09:00		Niha	OK
2	CAL2	CAL2	CAL	04/04/25 09:03		Niha	OK
3	CAL3	CAL3	CAL	04/04/25 09:06		Niha	OK
4	CAL4	CAL4	CAL	04/04/25 09:09		Niha	OK
5	CAL5	CAL5	CAL	04/04/25 09:12		Niha	OK
6	CAL6	CAL6	CAL	04/04/25 09:15		Niha	OK
7	ICV	ICV	ICV	04/04/25 09:18		Niha	OK
8	ICB	ICB	ICB	04/04/25 09:21		Niha	OK
9	CCV1	CCV1	CCV	04/04/25 09:24		Niha	OK
10	CCB1	CCB1	CCB	04/04/25 09:27		Niha	OK
11	LB135301BL	LB135301BL	MB	04/04/25 09:30		Niha	OK
12	LB135301BS	LB135301BS	LCS	04/04/25 09:33		Niha	OK
13	Q1715-01	WATER TREATMENT	SAM	04/04/25 09:36		Niha	OK
14	Q1715-01DUP	WATER TREATMENT	DUP	04/04/25 09:39		Niha	OK
15	Q1715-01MS	WATER TREATMENT	MS	04/04/25 09:42		Niha	OK
16	Q1715-01MSD	WATER TREATMENT	MSD	04/04/25 09:45		Niha	OK
17	CCV2	CCV2	CCV	04/04/25 09:48		Niha	OK
18	CCB2	CCB2	CCB	04/04/25 09:51		Niha	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB135357

Review By	rubina	Review On	4/9/2025 3:10:34 PM
Supervise By	Iwona	Supervise On	4/9/2025 3:11:38 PM
SubDirectory	LB135357	Test	Ammonia
STD. NAME	STD REF.#		
ICAL Standard	WP112644		
ICV Standard	WP112646		
CCV Standard	WP112645		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP112614		
Chk Standard	WP112537,WP111745,WP111385,WP111660		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPM	0.0PPM	CAL1	04/09/25 11:04		rubina	OK
2	0.1PPM	0.1PPM	CAL2	04/09/25 11:04		rubina	OK
3	0.2PPM	0.2PPM	CAL3	04/09/25 11:04		rubina	OK
4	0.4PPM	0.4PPM	CAL4	04/09/25 11:04		rubina	OK
5	1.0PPM	1.0PPM	CAL5	04/09/25 11:04		rubina	OK
6	1.3PPM	1.3PPM	CAL6	04/09/25 11:04		rubina	OK
7	2.0PPM	2.0PPM	CAL7	04/09/25 11:04		rubina	OK
8	ICV1	ICV1	ICV	04/09/25 11:35		rubina	OK
9	ICB1	ICB1	ICB	04/09/25 11:35		rubina	OK
10	CCV1	CCV1	CCV	04/09/25 11:35		rubina	OK
11	CCB1	CCB1	CCB	04/09/25 11:35		rubina	OK
12	RL	RL	SAM	04/09/25 11:35		rubina	OK
13	PB167481BL	PB167481BL	MB	04/09/25 11:35		rubina	OK
14	PB167481BS	PB167481BS	LCS	04/09/25 11:46		rubina	OK
15	Q1715-01	WATER TREATMENT	SAM	04/09/25 11:46		rubina	OK
16	Q1715-01DUP	WATER TREATMENT	DUP	04/09/25 11:46		rubina	OK
17	Q1715-01MS	WATER TREATMENT	MS	04/09/25 11:46		rubina	OK
18	Q1715-01MSD	WATER TREATMENT	MSD	04/09/25 11:46		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB135357

Review By	rubina	Review On	4/9/2025 3:10:34 PM
Supervise By	Iwona	Supervise On	4/9/2025 3:11:38 PM
SubDirectory	LB135357	Test	Ammonia
STD. NAME	STD REF.#		
ICAL Standard	WP112644		
ICV Standard	WP112646		
CCV Standard	WP112645		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP112614		
Chk Standard	WP112537,WP111745,WP111385,WP111660		

19	Q1717-01	EFFLUENT	SAM	04/09/25 11:46	High	rubina	Dilution
20	Q1717-05	INFLUENT	SAM	04/09/25 11:53	High	rubina	Dilution
21	CCV2	CCV2	CCV	04/09/25 11:53		rubina	OK
22	CCB2	CCB2	CCB	04/09/25 11:53		rubina	OK
23	Q1717-01DL	EFFLUENTDL	SAM	04/09/25 12:21	Report 10X	rubina	Confirms
24	Q1717-05DL	INFLUENTDL	SAM	04/09/25 12:21	Report 2X	rubina	Confirms
25	CCV3	CCV3	CCV	04/09/25 12:21		rubina	OK
26	CCB3	CCB3	CCB	04/09/25 12:21		rubina	OK