



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB135301

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.395	0.4	99	90-110	04/04/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.405	0.4	101	90-110	04/04/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.395	0.4	99	90-110	04/04/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB135357

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.95	1	95	90-110	04/09/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.96	1	96	90-110	04/09/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	04/09/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.99	1	99	90-110	04/09/2025