

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1715
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1715-01
Client ID:	WATER TREATMENT DISCHARGEMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	0.41		0.052	J	0.4	1	88		04/04/2025
Ammonia as N	mg/L	75-125	0.96		0.040	J	1	1	92		04/09/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1715
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q1715-01
Client ID:	WATER TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	0.41		0.052	J	0.4	1	88		04/04/2025
Ammonia as N	mg/L	75-125	0.96		0.040	J	1	1	92		04/09/2025