

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1715
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB135301

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135301BS							
Residual Chlorine	mg/L	0.4	0.40		99	1	90-110	04/04/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1715
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB135357

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167481BS							
Ammonia as N	mg/L	1	0.96		96	1	90-110	04/09/2025