



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1715 **Order ID:** Q1715
Client: VERINA CONSULTING GROUP, LLC **Project ID:** Rotor Clip NJ WTD - 2025

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : WATER TREATMENT DISCHARGE								
Q1715-01	WATER TREATMENT DISCHA	Water	Chromium	465		0.66	5.00	ug/L
Q1715-01	WATER TREATMENT DISCHA	Water	Copper	10.5		7.07	10.0	ug/L
Q1715-01	WATER TREATMENT DISCHA	Water	Nickel	18.3	J	0.85	20.0	ug/L
Q1715-01	WATER TREATMENT DISCHA	Water	Zinc	612		1.75	20.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	VERINA CONSULTING GROUP, LLC	Date Collected:	04/03/25
Project:	Rotor Clip NJ WTD - 2025	Date Received:	04/03/25
Client Sample ID:	WATER TREATMENT DISCHARGE	SDG No.:	Q1715
Lab Sample ID:	Q1715-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-47-3	Chromium	465		1	0.66	5.00	ug/L	04/07/25 11:10	04/10/25 11:42	SW6010	SW3010
7440-50-8	Copper	10.5		1	7.07	10.0	ug/L	04/07/25 11:10	04/10/25 11:42	SW6010	SW3010
7440-02-0	Nickel	18.3	J	1	0.85	20.0	ug/L	04/07/25 11:10	04/10/25 11:42	SW6010	SW3010
7440-66-6	Zinc	612		1	1.75	20.0	ug/L	04/07/25 11:10	04/10/25 11:42	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	10:29	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC		SDG No.: Q1715							
Contract: VERI01	Lab Code: CHEM	Case No.: Q1715	SAS No.: Q1715						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	11:21	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
CCB02	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	12:12	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
CCB03	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	13:04	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
CCB04	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	13:55	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
CCB05	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	14:49	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	15:48	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
CCB07	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	17:00	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
CCB08	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	18:00	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
CCB09	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	18:23	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167483BL	WATER			Batch Number:	PB167483		Prep Date:	04/07/2025	
	Chromium	5.00	<5.00	U	5.00	P	04/10/2025	11:26	LB135373
	Copper	10.0	<10.0	U	10.0	P	04/10/2025	11:26	LB135373
	Nickel	20.0	<20.0	U	20.0	P	04/10/2025	11:26	LB135373
	Zinc	20.0	<20.0	U	20.0	P	04/10/2025	11:26	LB135373



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	505	520	97	90 - 110	P	04/10/2025	10:09	LB135373
	Copper	517	510	102	90 - 110	P	04/10/2025	10:09	LB135373
	Nickel	497	530	94	90 - 110	P	04/10/2025	10:09	LB135373
	Zinc	996	1000	100	90 - 110	P	04/10/2025	10:09	LB135373

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	9.56	10.0	96	80 - 120	P	04/10/2025	10:25	LB135373
	Copper	21.9	20.0	110	80 - 120	P	04/10/2025	10:25	LB135373
	Nickel	37.7	40.0	94	80 - 120	P	04/10/2025	10:25	LB135373
	Zinc	39.6	40.0	99	80 - 120	P	04/10/2025	10:25	LB135373

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC SDG No.: Q1715
 Contract: VERI01 Lab Code: CHEM Case No.: Q1715 SAS No.: Q1715
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	1010	1000	101	90 - 110	P	04/10/2025	11:00	LB135373
	Copper	1260	1250	101	90 - 110	P	04/10/2025	11:00	LB135373
	Nickel	2460	2500	98	90 - 110	P	04/10/2025	11:00	LB135373
	Zinc	2540	2500	102	90 - 110	P	04/10/2025	11:00	LB135373
CCV02	Chromium	1020	1000	102	90 - 110	P	04/10/2025	12:08	LB135373
	Copper	1270	1250	102	90 - 110	P	04/10/2025	12:08	LB135373
	Nickel	2480	2500	99	90 - 110	P	04/10/2025	12:08	LB135373
	Zinc	2530	2500	101	90 - 110	P	04/10/2025	12:08	LB135373
CCV03	Chromium	1020	1000	102	90 - 110	P	04/10/2025	12:59	LB135373
	Copper	1290	1250	103	90 - 110	P	04/10/2025	12:59	LB135373
	Nickel	2490	2500	100	90 - 110	P	04/10/2025	12:59	LB135373
	Zinc	2560	2500	102	90 - 110	P	04/10/2025	12:59	LB135373
CCV04	Chromium	1020	1000	102	90 - 110	P	04/10/2025	13:51	LB135373
	Copper	1290	1250	103	90 - 110	P	04/10/2025	13:51	LB135373
	Nickel	2500	2500	100	90 - 110	P	04/10/2025	13:51	LB135373
	Zinc	2570	2500	103	90 - 110	P	04/10/2025	13:51	LB135373
CCV05	Chromium	1020	1000	102	90 - 110	P	04/10/2025	14:44	LB135373
	Copper	1280	1250	103	90 - 110	P	04/10/2025	14:44	LB135373
	Nickel	2480	2500	99	90 - 110	P	04/10/2025	14:44	LB135373
	Zinc	2550	2500	102	90 - 110	P	04/10/2025	14:44	LB135373
CCV06	Chromium	1010	1000	101	90 - 110	P	04/10/2025	15:44	LB135373
	Copper	1270	1250	102	90 - 110	P	04/10/2025	15:44	LB135373
	Nickel	2470	2500	99	90 - 110	P	04/10/2025	15:44	LB135373
	Zinc	2520	2500	101	90 - 110	P	04/10/2025	15:44	LB135373
CCV07	Chromium	1020	1000	102	90 - 110	P	04/10/2025	16:56	LB135373
	Copper	1290	1250	103	90 - 110	P	04/10/2025	16:56	LB135373
	Nickel	2500	2500	100	90 - 110	P	04/10/2025	16:56	LB135373
	Zinc	2560	2500	102	90 - 110	P	04/10/2025	16:56	LB135373
CCV08	Chromium	1030	1000	103	90 - 110	P	04/10/2025	17:56	LB135373
	Copper	1280	1250	103	90 - 110	P	04/10/2025	17:56	LB135373
	Nickel	2500	2500	100	90 - 110	P	04/10/2025	17:56	LB135373
	Zinc	2550	2500	102	90 - 110	P	04/10/2025	17:56	LB135373
CCV09	Chromium	1000	1000	100	90 - 110	P	04/10/2025	18:19	LB135373
	Copper	1290	1250	103	90 - 110	P	04/10/2025	18:19	LB135373

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV09	Nickel	2480	2500	99	90 - 110	P	04/10/2025	18:19	LB135373
	Zinc	2500	2500	100	90 - 110	P	04/10/2025	18:19	LB135373



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	9.42	10.0	94	40 - 160	P	04/10/2025	10:33	LB135373
	Copper	21.9	20.0	109	40 - 160	P	04/10/2025	10:33	LB135373
	Nickel	38.0	40.0	95	40 - 160	P	04/10/2025	10:33	LB135373
	Zinc	39.6	40.0	99	40 - 160	P	04/10/2025	10:33	LB135373

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	55.8	52.0	107	42	62	04/10/2025	10:37	LB135373
	Copper	-4.96	2.0	248	-18	22	04/10/2025	10:37	LB135373
	Nickel	2.46	2.0	123	-38	42	04/10/2025	10:37	LB135373
	Zinc	2.74			-40	40	04/10/2025	10:37	LB135373
ICSAB01	Chromium	553	542	102	460	624	04/10/2025	10:42	LB135373
	Copper	500	511	98	434	588	04/10/2025	10:42	LB135373
	Nickel	995	954	104	810	1100	04/10/2025	10:42	LB135373
	Zinc	1060	952	111	809	1095	04/10/2025	10:42	LB135373



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q1715
contract: VERI01 **lab code:** CHEM **case no.:** Q1715 **sas no.:** Q1715
matrix: Water **sample id:** Q1715-01 **client id:** WATER TREATMENT DISCHARGEMS
Percent Solids for Sample: NA **Spiked ID:** Q1715-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	619		465		200	77		P
Copper	ug/L	75 - 125	130		10.5		150	80		P
Nickel	ug/L	75 - 125	231		18.3	J	250	85		P
Zinc	ug/L	75 - 125	684		612		100	72		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q1715
contract: VERI01 **lab code:** CHEM **case no.:** Q1715 **sas no.:** Q1715
matrix: Water **sample id:** Q1715-01 **client id:** WATER TREATMENT DISCHARGEMSD
Percent Solids for Sample: NA **Spiked ID:** Q1715-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	620		465		200	78		P
Copper	ug/L	75 - 125	129		10.5		150	79		P
Nickel	ug/L	75 - 125	231		18.3	J	250	85		P
Zinc	ug/L	75 - 125	683		612		100	71		P

Metals
- 5b -

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Case No.: Q1715 **SAS No.:** Q1715

Matrix:

Level: LOW

Client ID:

Sample ID: **Spiked ID:**

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

A
B
C
D
E
F
G
H
I
J

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715
Matrix: Water **Sample ID:** Q1715-01 **Client ID:** WATER TREATMENT DISCHARGEDUP
Percent Solids for Sample: NA **Duplicate ID** Q1715-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	465		474		2		P
Copper	ug/L	20	10.5		10.6		1		P
Nickel	ug/L	20	18.3	J	18.6	J	2		P
Zinc	ug/L	20	612		622		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715
Matrix: Water **Sample ID:** Q1715-01MS **Client ID:** WATER TREATMENT DISCHARGEMSD
Percent Solids for Sample: NA **Duplicate ID** Q1715-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	619		620		0		P
Copper	ug/L	20	130		129		1		P
Nickel	ug/L	20	231		231		0		P
Zinc	ug/L	20	684		683		0		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Case No.: Q1715

SAS No.: Q1715

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167483BS							
Chromium	ug/L	200	193		96	80 - 120	P
Copper	ug/L	150	150		100	80 - 120	P
Nickel	ug/L	250	235		94	80 - 120	P
Zinc	ug/L	100	97.4		97	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WATER TREATMENT DISCHARGEL

Lab Name: Chemtech Consulting Group **Contract:** VERI01
Lab Code: CHEM **Lb No.:** lb135373 **Lab Sample ID :** Q1715-01L **SDG No.:** Q1715
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence		Q	M
Chromium	465		486		5			P
Copper	10.5		50.0	U	100.0			P
Nickel	18.3	J	18.0	J	1			P
Zinc	612		607		1			P

metals
- 14 -
ANALYSIS RUN LOG

Client: VERINA CONSULTING GROUP, LLC

Contract: VERI01

Lab code: CHEM

Case no.: Q1715

Sas no.: Q1715

Sdg no.: Q1715

Instrument id number:

Method:

Run number: LB135373

Start date: 04/10/2025

End date: 04/10/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	0930	Cr,Cu,Ni,Zn
S1	S1	1	0935	Cr,Cu,Ni,Zn
S2	S2	1	0939	Cr,Cu,Ni,Zn
S3	S3	1	0943	Cr,Cu,Ni,Zn
S4	S4	1	0947	Cr,Cu,Ni,Zn
S5	S5	1	0951	Cr,Cu,Ni,Zn
ICV01	ICV01	1	1009	Cr,Cu,Ni,Zn
LLICV01	LLICV01	1	1025	Cr,Cu,Ni,Zn
ICB01	ICB01	1	1029	Cr,Cu,Ni,Zn
CRI01	CRI01	1	1033	Cr,Cu,Ni,Zn
ICSA01	ICSA01	1	1037	Cr,Cu,Ni,Zn
ICSAB01	ICSAB01	1	1042	Cr,Cu,Ni,Zn
CCV01	CCV01	1	1100	Cr,Cu,Ni,Zn
CCB01	CCB01	1	1121	Cr,Cu,Ni,Zn
PB167483BL	PB167483BL	1	1126	Cr,Cu,Ni,Zn
PB167483BS	PB167483BS	1	1130	Cr,Cu,Ni,Zn
Q1715-01	WATER TREATMENT DISCHA	1	1142	Cr,Cu,Ni,Zn
CCV02	CCV02	1	1208	Cr,Cu,Ni,Zn
CCB02	CCB02	1	1212	Cr,Cu,Ni,Zn
CCV03	CCV03	1	1259	Cr,Cu,Ni,Zn
CCB03	CCB03	1	1304	Cr,Cu,Ni,Zn
Q1715-01DUP	WATER TREATMENT DISCHA	1	1333	Cr,Cu,Ni,Zn
Q1715-01L	WATER TREATMENT DISCHA	5	1338	Cr,Cu,Ni,Zn
CCV04	CCV04	1	1351	Cr,Cu,Ni,Zn
CCB04	CCB04	1	1355	Cr,Cu,Ni,Zn
Q1715-01MS	WATER TREATMENT DISCHA	1	1359	Cr,Cu,Ni,Zn
Q1715-01MSD	WATER TREATMENT DISCHA	1	1407	Cr,Cu,Ni,Zn
CCV05	CCV05	1	1444	Cr,Cu,Ni,Zn
CCB05	CCB05	1	1449	Cr,Cu,Ni,Zn
CCV06	CCV06	1	1544	Cr,Cu,Ni,Zn
CCB06	CCB06	1	1548	Cr,Cu,Ni,Zn
CCV07	CCV07	1	1656	Cr,Cu,Ni,Zn
CCB07	CCB07	1	1700	Cr,Cu,Ni,Zn
CCV08	CCV08	1	1756	Cr,Cu,Ni,Zn
CCB08	CCB08	1	1800	Cr,Cu,Ni,Zn
CCV09	CCV09	1	1819	Cr,Cu,Ni,Zn
CCB09	CCB09	1	1823	Cr,Cu,Ni,Zn



METAL PREPARATION & INSTRUMENT DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Case No.: Q1715

SAS No.: Q1715

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Case No.: Q1715

SAS No.: Q1715

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789
8900, Fax : 908 789 8922

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Case No.: Q1715

SAS No.: Q1715

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Case No.: Q1715

SAS No.: Q1715

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Case No.: Q1715

SAS No.: Q1715

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000



LAB CHRONICLE

OrderID:	Q1715	OrderDate:	4/3/2025 12:01:00 PM
Client:	VERINA CONSULTING GROUP, LLC	Project:	Rotor Clip NJ WTD - 2025
Contact:	Michael Valenzi	Location:	F11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1715-01	WATER TREATMENT DISCHARGE	Water	Metals Group4	6010D	04/03/25	04/07/25	04/10/25	04/03/25



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Method: _____

Case No.: Q1715 **SAS No.:** Q1715

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB167483							
PB167483BL	PB167483BL	MB	WATER	04/07/2025	50.0	25.0	
PB167483BS	PB167483BS	LCS	WATER	04/07/2025	50.0	25.0	
Q1715-01	WATER TREATMENT DISCHARGE	SAM	WATER	04/07/2025	50.0	25.0	
Q1715-01DUP	WATER TREATMENT DISCHARGEDUP	DUP	WATER	04/07/2025	50.0	25.0	
Q1715-01MS	WATER TREATMENT DISCHARGEMS	MS	WATER	04/07/2025	50.0	25.0	
Q1715-01MSD	WATER TREATMENT DISCHARGEMSD	MSD	WATER	04/07/2025	50.0	25.0	

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135373

Review By	jaswal	Review On	4/11/2025 8:35:52 AM
Supervise By	mohan	Supervise On	4/11/2025 8:55:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85016,MP85017,MP85018,MP85019,MP85020,MP85022		
ICV Standard	MP85023		
CCV Standard	MP85026		
ICSA Standard	MP85024,MP85025		
CRI Standard	MP85022		
LCS Standard			
Chk Standard	MP85030,MP85031		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	04/10/25 09:30		Kareem	OK
2	S1	S1	CAL2	04/10/25 09:35		Kareem	OK
3	S2	S2	CAL3	04/10/25 09:39		Kareem	OK
4	S3	S3	CAL4	04/10/25 09:43		Kareem	OK
5	S4	S4	CAL5	04/10/25 09:47		Kareem	OK
6	S5	S5	CAL6	04/10/25 09:51		Kareem	OK
7	ICV01	ICV01	ICV	04/10/25 10:09	ICV01 Fail For Be,Cd,Ca,Co,Mg,Mn,Ni,V (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	04/10/25 10:25		Kareem	OK
9	ICB01	ICB01	ICB	04/10/25 10:29		Kareem	OK
10	CRI01	CRI01	CRDL	04/10/25 10:33		Kareem	OK
11	ICSA01	ICSA01	ICSA	04/10/25 10:37		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	04/10/25 10:42		Kareem	OK
13	ICSADL	ICSADL	ICSA	04/10/25 10:46		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	04/10/25 10:50		Kareem	OK
15	CCV01	CCV01	CCV	04/10/25 11:00		Kareem	OK
16	CCB01	CCB01	CCB	04/10/25 11:21		Kareem	OK
17	PB167483BL	PB167483BL	MB	04/10/25 11:26		Kareem	OK
18	PB167483BS	PB167483BS	LCS	04/10/25 11:30		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135373

Review By	jaswal	Review On	4/11/2025 8:35:52 AM
Supervise By	mohan	Supervise On	4/11/2025 8:55:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85016,MP85017,MP85018,MP85019,MP85020,MP85022		
ICV Standard	MP85023		
CCV Standard	MP85026		
ICSA Standard	MP85024,MP85025		
CRI Standard	MP85022		
LCS Standard			
Chk Standard	MP85030,MP85031		

19	PB167497BL	PB167497BL	MB	04/10/25 11:34		Kareem	OK
20	PB167497BS	PB167497BS	LCS	04/10/25 11:38		Kareem	OK
21	Q1715-01	WATER TREATMENT	SAM	04/10/25 11:42		Kareem	OK
22	Q1745-01	IB-6A-WC	SAM	04/10/25 11:47		Kareem	OK
23	Q1745-09	IB-6.5-WC	SAM	04/10/25 11:51		Kareem	OK
24	Q1748-01	IB-1.5-WC	SAM	04/10/25 11:55		Kareem	OK
25	Q1748-05	IB-2A-WC	SAM	04/10/25 11:59		Kareem	OK
26	Q1749-01	TP-14	SAM	04/10/25 12:03		Kareem	OK
27	CCV02	CCV02	CCV	04/10/25 12:08		Kareem	OK
28	CCB02	CCB02	CCB	04/10/25 12:12		Kareem	OK
29	Q1732-01	TT-8	SAM	04/10/25 12:17		Kareem	OK
30	Q1733-01	ETGI-328	SAM	04/10/25 12:21		Kareem	OK
31	Q1740-01	TP-20	SAM	04/10/25 12:25		Kareem	OK
32	Q1743-01	TP-16	SAM	04/10/25 12:29		Kareem	OK
33	Q1739-01	WC-LIQUID-20250404	SAM	04/10/25 12:33		Kareem	OK
34	Q1748-04	IB-1.5-WC	SAM	04/10/25 12:38		Kareem	OK
35	Q1748-08	IB-2A-WC	SAM	04/10/25 12:42		Kareem	OK
36	Q1754-02	TP-1	SAM	04/10/25 12:46		Kareem	OK
37	Q1754-04	TP-1-CONCRETE	SAM	04/10/25 12:51		Kareem	OK
38	Q1754-01	TP-1	SAM	04/10/25 12:55		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135373

Review By	jaswal	Review On	4/11/2025 8:35:52 AM
Supervise By	mohan	Supervise On	4/11/2025 8:55:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85016,MP85017,MP85018,MP85019,MP85020,MP85022		
ICV Standard	MP85023		
CCV Standard	MP85026		
ICSA Standard	MP85024,MP85025		
CRI Standard	MP85022		
LCS Standard			
Chk Standard	MP85030,MP85031		

39	CCV03	CCV03	CCV	04/10/25 12:59		Kareem	OK
40	CCB03	CCB03	CCB	04/10/25 13:04		Kareem	OK
41	Q1754-03	TP-1-CONCRETE	SAM	04/10/25 13:07		Kareem	OK
42	Q1754-03DUP	TP-1-CONCRETEDUP	DUP	04/10/25 13:12		Kareem	OK
43	Q1754-03L	TP-1-CONCRETEL	SD	04/10/25 13:16		Kareem	OK
44	Q1754-03A	TP-1-CONCRETEA	PS	04/10/25 13:29		Kareem	OK
45	Q1715-01DUP	WATER TREATMENT	DUP	04/10/25 13:33		Kareem	OK
46	Q1715-01L	WATER TREATMENT	SD	04/10/25 13:38		Kareem	OK
47	Q1745-04	IB-6A-WC	SAM	04/10/25 13:42		Kareem	OK
48	Q1745-12	IB-6.5-WC	SAM	04/10/25 13:47		Kareem	OK
49	CCV04	CCV04	CCV	04/10/25 13:51		Kareem	OK
50	CCB04	CCB04	CCB	04/10/25 13:55		Kareem	OK
51	Q1715-01MS	WATER TREATMENT	MS	04/10/25 13:59		Kareem	OK
52	Q1715-01MSD	WATER TREATMENT	MSD	04/10/25 14:07		Kareem	OK
53	Q1715-01A	WATER TREATMENT	PS	04/10/25 14:11		Kareem	OK
54	Q1744-01	B-158-SB01	SAM	04/10/25 14:15		Kareem	OK
55	Q1744-03	B-158-SB02	SAM	04/10/25 14:19		Kareem	OK
56	Q1754-03MS	TP-1-CONCRETEMS	MS	04/10/25 14:23		Kareem	OK
57	Q1754-03MSD	TP-1-CONCRETEMS	MSD	04/10/25 14:27		Kareem	OK
58	Q1745-09DUP	IB-6.5-WCDUP	DUP	04/10/25 14:32		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135373

Review By	jaswal	Review On	4/11/2025 8:35:52 AM
Supervise By	mohan	Supervise On	4/11/2025 8:55:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85016,MP85017,MP85018,MP85019,MP85020,MP85022		
ICV Standard	MP85023		
CCV Standard	MP85026		
ICSA Standard	MP85024,MP85025		
CRI Standard	MP85022		
LCS Standard			
Chk Standard	MP85030,MP85031		

59	Q1745-09L	IB-6.5-WCL	SD	04/10/25 14:36		Kareem	OK
60	CCV05	CCV05	CCV	04/10/25 14:44		Kareem	OK
61	CCB05	CCB05	CCB	04/10/25 14:49		Kareem	OK
62	Q1745-09A	IB-6.5-WCA	PS	04/10/25 14:57		Kareem	OK
63	Q1745-09MS	IB-6.5-WCMS	MS	04/10/25 15:05		Kareem	OK
64	Q1745-09MSD	IB-6.5-WCMSD	MSD	04/10/25 15:09		Kareem	OK
65	PBW04	PBW04	MB	04/10/25 15:14		Kareem	OK
66	LCSSW04	LCSSW04	LCS	04/10/25 15:18		Kareem	OK
67	PBS04	PBS04	MB	04/10/25 15:23		Kareem	OK
68	LCSS04	LCSS04	LCS	04/10/25 15:27		Kareem	OK
69	PBW05	PBW05	MB	04/10/25 15:31		Kareem	OK
70	PBS05	PBS05	MB	04/10/25 15:35		Kareem	OK
71	WATER MDL-4	WATER MDL-4	SAM	04/10/25 15:40		Kareem	OK
72	CCV06	CCV06	CCV	04/10/25 15:44		Kareem	OK
73	CCB06	CCB06	CCB	04/10/25 15:48		Kareem	OK
74	WATER MDL-5	WATER MDL-5	SAM	04/10/25 15:52		Kareem	OK
75	SOIL MDL-4	SOIL MDL-4	SAM	04/10/25 16:09		Kareem	OK
76	Q1746-01	B-149-SB01	SAM	04/10/25 16:17		Kareem	OK
77	Q1746-03	B-149-SB02	SAM	04/10/25 16:22		Kareem	OK
78	PB167490BL	PB167490BL	MB	04/10/25 16:26		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135373

Review By	jaswal	Review On	4/11/2025 8:35:52 AM
Supervise By	mohan	Supervise On	4/11/2025 8:55:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85016,MP85017,MP85018,MP85019,MP85020,MP85022		
ICV Standard	MP85023		
CCV Standard	MP85026		
ICSA Standard	MP85024,MP85025		
CRI Standard	MP85022		
LCS Standard			
Chk Standard	MP85030,MP85031		

79	PB167490BS	PB167490BS	LCS	04/10/25 16:39		Kareem	OK
80	PB167517TB	PB167517TB	MB	04/10/25 16:43		Kareem	OK
81	Q1744-02	B-158-SB01	SAM	04/10/25 16:47		Kareem	OK
82	Q1744-04	B-158-SB02	SAM	04/10/25 16:52	Confirm TCLP for Ba	Kareem	OK
83	CCV07	CCV07	CCV	04/10/25 16:56		Kareem	OK
84	CCB07	CCB07	CCB	04/10/25 17:00		Kareem	OK
85	Q1780-01	001 WILLETS PT BLV	SAM	04/10/25 17:05		Kareem	OK
86	Q1780-02	002 35TH AVE (MAR)	SAM	04/10/25 17:09		Kareem	OK
87	Q1780-02DUP	002 35TH AVE (MAR)	DUP	04/10/25 17:13		Kareem	OK
88	Q1780-02L	002 35TH AVE (MAR)	SD	04/10/25 17:18		Kareem	OK
89	Q1780-02MS	002 35TH AVE (MAR)	MS	04/10/25 17:22		Kareem	OK
90	Q1780-02MSD	002 35TH AVE (MAR)	MSD	04/10/25 17:26		Kareem	OK
91	Q1780-02A	002 35TH AVE (MAR)	PS	04/10/25 17:30		Kareem	OK
92	SOIL MDL-5	SOIL MDL-5	SAM	04/10/25 17:34		Kareem	OK
93	PB167554BL	PB167554BL	MB	04/10/25 17:39		Kareem	OK
94	PB167554BS	PB167554BS	LCS	04/10/25 17:43		Kareem	OK
95	CCV08	CCV08	CCV	04/10/25 17:56		Kareem	OK
96	CCB08	CCB08	CCB	04/10/25 18:00		Kareem	OK
97	LR1	LR1	HIGH STD	04/10/25 18:05		Kareem	OK
98	LR2	LR2	HIGH STD	04/10/25 18:10		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135373

Review By	jaswal	Review On	4/11/2025 8:35:52 AM
Supervise By	mohan	Supervise On	4/11/2025 8:55:22 AM

STD. NAME	STD REF.#
ICAL Standard	MP85016,MP85017,MP85018,MP85019,MP85020,MP85022
ICV Standard	MP85023
CCV Standard	MP85026
ICSA Standard	MP85024,MP85025
CRI Standard	MP85022
LCS Standard	
Chk Standard	MP85030,MP85031

99	CCV09	CCV09	CCV	04/10/25 18:19		Kareem	OK
100	CCB09	CCB09	CCB	04/10/25 18:23		Kareem	OK

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 04/07/2025 Time : 11:10 Temp : 96 °C

End Digest Date: 04/07/2025 Time : 14:15 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: 

Supervisor Signature: 


Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6003
LFS-1	0.25	M6012
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP85156
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK # 1 CELL 50 Temp :96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/07/25 15:15	SIB. met. dig	
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB167483BL	PBW483	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	1
PB167483BS	LCS483	<2	50	25	Colorless	Colorless	Clear	Clear	M6003,M6012	2
Q1604-05	FRAC-TANK-N45878	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
Q1715-01	WATER TREATMENT DISCHARGE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
Q1715-01MS	WATER TREATMENT DISCHARGE	<2	50	25	Colorless	Colorless	Clear	Clear	M6003,M6012	6
Q1715-01MSD	WATER TREATMENT DISCHARGE	<2	50	25	Colorless	Colorless	Clear	Clear	M6003,M6012	7
Q1715-01DUP	WATER TREATMENT DISCHARGE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	5
Q1739-01	WC-LIQUID-20250404	<2	50	50	Colorless	Colorless	Clear	Clear	N/A	8