

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1715
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1715 **SAS No.:** Q1715
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	55.8	52.0	107	42	62	04/10/2025	10:37	LB135373
	Copper	-4.96	2.0	248	-18	22	04/10/2025	10:37	LB135373
	Nickel	2.46	2.0	123	-38	42	04/10/2025	10:37	LB135373
	Zinc	2.74			-40	40	04/10/2025	10:37	LB135373
ICSAB01	Chromium	553	542	102	460	624	04/10/2025	10:42	LB135373
	Copper	500	511	98	434	588	04/10/2025	10:42	LB135373
	Nickel	995	954	104	810	1100	04/10/2025	10:42	LB135373
	Zinc	1060	952	111	809	1095	04/10/2025	10:42	LB135373