

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1715

Contract: VERI01

Lab Code: CHEM

Case No.: Q1715

SAS No.: Q1715

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167483BS							
Chromium	ug/L	200	193		96	80 - 120	P
Copper	ug/L	150	150		100	80 - 120	P
Nickel	ug/L	250	235		94	80 - 120	P
Zinc	ug/L	100	97.4		97	80 - 120	P