

**Metals**

-9 -

**ICP SERIAL DILUTIONS**

**SAMPLE NO.**

WATER TREATMENT DISCHARGE

**Lab Name:** Chemtech Consulting Group

**Contract:** VERI01

**Lab Code:** CHEM **Lb No.:** lb135373 **Lab Sample ID :** Q1715-01L **SDG No.:** Q1715

**Matrix (soil/water):** Water **Level (low/med):** LOW

**Concentration Units:** ug/L

| Analyte  | Initial Sample<br>Result (I) |   | Serial Dilution<br>Result (S) |   | % Differ-<br>ence |  | Q | M |
|----------|------------------------------|---|-------------------------------|---|-------------------|--|---|---|
|          | C                            |   | C                             |   |                   |  |   |   |
| Chromium | 465                          |   | 486                           |   | 5                 |  |   | P |
| Copper   | 10.5                         |   | 50.0                          | U | 100.0             |  |   | P |
| Nickel   | 18.3                         | J | 18.0                          | J | 1                 |  |   | P |
| Zinc     | 612                          |   | 607                           |   | 1                 |  |   | P |