

SOP ID : MSM4500-NH3 B,G-Ammonia-17

SDG No : N/A

Start Digest Date: 04/09/2025 Time : 10:00 Temp : 150 °C

Matrix : WATER

End Digest Date: 04/09/2025 Time : 11:00 Temp : 160 °C

Pipette ID : WC

Balance ID : N/A

Hood ID : HOOD#2

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : WC-DIST-BLOCK-1

Filter paper ID : N/A

Prep Technician Signature: RM

Weigh By : N/A

pH Meter ID : N/A

Supervisor Signature: 12

Standardized Name	MLS USED	STD REF. # FROM LOG
LCSW	1.0ML	WP112614
MS/MSD SPIKE SOL.	1.0ML	WP112613
PBW	50.0ML	W3112
RL CHECK	0.1ML	WP112613
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP111325
NAOH 6N	0.5-2.0ML	WP111318
H2SO4 0.04N	5.0ML	WP110335
pH strip-Ammonia	N/A	W3133
KI-starch paper	N/A	W3155
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT WP111604. Due to bad matrix and client history 1ML was taken as an initial volume for P1717-01 and P1717-05.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/09/2025 11:10	RM (WC)	RM (WC)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/Nitrite	Comment	Prep Pos
PB167481BL	PBW481	50	50	N/A	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB167481BS	LCS481	50	50	N/A	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1715-01DUP	WATER TREATMENT DISCHARGEDUP	50	50	N/A	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1715-01MS	WATER TREATMENT DISCHARGEMS	50	50	N/A	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1715-01MSD	WATER TREATMENT DISCHARGEMSD	50	50	N/A	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1715-01	WATER TREATMENT DISCHARGE	50	50	N/A	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1717-01	EFFLUENT	1	50	N/A	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1717-05	INFLUENT	1	50	N/A	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A