

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

RunNo.: LB135290

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.505	0.50	101	90-110	04/03/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.499	0.5	100	90-110	04/03/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.502	0.5	100	90-110	04/03/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

RunNo.: LB135292

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.499	0.50	100	90-110	04/03/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.496	0.50	99	90-110	04/03/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.505	0.50	101	90-110	04/03/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

RunNo.: LB135357

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.95	1	95	90-110	04/09/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.96	1	96	90-110	04/09/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	04/09/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.99	1	99	90-110	04/09/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

RunNo.: LB135357

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------