

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

Run No.: LB135290

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135290BS							
Orthophosphate as P	mg/L	0.5	0.51		102	1	90-110	04/03/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

Run No.: LB135306

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135306BS							
BOD5	mg/L	198	182		92	1	84.6-115.4	04/04/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

Run No.: LB135334

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135334BS							
TSS	mg/L	550	533		97	1	90-110	04/08/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

Run No.: LB135353

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135353BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	04/09/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

Run No.: LB135292

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167453BS							
Phosphorus, Total	mg/L	0.50	0.50		100	1	90-110	04/03/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1717

Project: Pre Treatment Plant 2025

Run No.: LB135357

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167481BS							
Ammonia as N	mg/L	1	0.96		96	1	90-110	04/09/2025