

Duplicate Sample Summary

Lab Sample Id :	Q1725-04DUP	Q1725-04
Client Id :	AA-1-03312025DUP	AA-1-03312025
DF :	1	1
Datafile :	VL042232.D	VL042231.D
Anal Date & Time :	04/07/2025 13:00	04/07/2025 12:27

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.49	0.53	7.8
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0.12	0.13	8
Acetone	5.2	5.1	1.9
Allyl Chloride	0	0	0
Benzene	0.13	0.14	7.4
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.06	0.07	15.4

Duplicate Sample Summary

Lab Sample Id :	Q1725-04DUP	Q1725-04
Client Id :	AA-1-03312025DUP	AA-1-03312025
DF :	1	1
Datafile :	VL042232.D	VL042231.D
Anal Date & Time :	04/07/2025 13:00	04/07/2025 12:27

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.56	0.58	3.5
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.4	0.4	0
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0.12	0.11	8.7
Hexachloro-1,3-Butadiene	0	0	0
Hexane	0.33	0.33	0
m/p-Xylene	0.22	0.22	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.48	0.4	18.2
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0	0	0
Tetrahydrofuran	0	0	0
Toluene	0.59	0.61	3.3
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.21	0.21	0
Vinyl Chloride	0	0	0